

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG NO.: P5026

Continuing Calib Date: 12/02/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 10:18 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	5.53	5.53	5.43	5.63	0.00
Aroclor-1242-2	(2)	5.55	5.55	5.45	5.65	0.00
Aroclor-1242-3	(3)	5.61	5.61	5.51	5.71	0.00
Aroclor-1242-4	(4)	5.71	5.71	5.61	5.81	0.00
Aroclor-1242-5	(5)	6.45	6.45	6.35	6.55	0.00
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.10	10.09	9.99	10.19	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG NO.: P5026

Continuing Calib Date: 12/02/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 10:18 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1242-2 (2)	4.74	4.74	4.64	4.84	0.00
Aroclor-1242-3 (3)	4.91	4.92	4.82	5.02	0.01
Aroclor-1242-4 (4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1242-5 (5)	5.52	5.52	5.42	5.62	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.63	8.63	8.53	8.73	0.00