

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG NO.: P5026

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 12/02/2024

Lab Sample No.: AR1660CCC500 Data File : PO108275.D Time Analyzed: 10:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.530	5.425	5.625	496.620	500.000	-0.7
Aroclor-1016-2	5.552	5.447	5.647	500.300	500.000	0.1
Aroclor-1016-3	5.615	5.510	5.710	493.010	500.000	-1.4
Aroclor-1016-4	5.712	5.607	5.807	504.490	500.000	0.9
Aroclor-1016-5	6.008	5.904	6.104	496.090	500.000	-0.8
Aroclor-1260-1	7.143	7.037	7.237	451.790	500.000	-9.6
Aroclor-1260-2	7.400	7.295	7.495	512.580	500.000	2.5
Aroclor-1260-3	7.764	7.658	7.858	509.460	500.000	1.9
Aroclor-1260-4	7.988	7.884	8.084	542.430	500.000	8.5
Aroclor-1260-5	8.304	8.198	8.398	544.180	500.000	8.8
Decachlorobiphenyl	10.100	9.991	10.191	52.240	50.000	4.5
Tetrachloro-m-xylene	4.374	4.270	4.470	52.880	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG NO.: P5026

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 12/02/2024

Lab Sample No.: AR1660CCC500 Data File : PO108275.D Time Analyzed: 10:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.620	4.820	546.000	500.000	9.2
Aroclor-1016-2	4.738	4.639	4.839	552.240	500.000	10.4
Aroclor-1016-3	4.914	4.815	5.015	556.490	500.000	11.3
Aroclor-1016-4	4.956	4.856	5.056	559.120	500.000	11.8
Aroclor-1016-5	5.169	5.070	5.270	546.670	500.000	9.3
Aroclor-1260-1	6.200	6.100	6.300	529.460	500.000	5.9
Aroclor-1260-2	6.388	6.288	6.488	526.080	500.000	5.2
Aroclor-1260-3	6.541	6.441	6.641	527.020	500.000	5.4
Aroclor-1260-4	7.012	6.912	7.112	520.110	500.000	4.0
Aroclor-1260-5	7.253	7.154	7.354	510.030	500.000	2.0
Decachlorobiphenyl	8.634	8.535	8.735	44.990	50.000	-10.0
Tetrachloro-m-xylene	3.639	3.539	3.739	57.590	50.000	15.2