

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG NO.: P5026

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 12/02/2024

Lab Sample No.: AR1660CCC500 Data File : PO108290.D Time Analyzed: 15:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.425	5.625	487.990	500.000	-2.4
Aroclor-1016-2	5.548	5.447	5.647	481.260	500.000	-3.7
Aroclor-1016-3	5.610	5.510	5.710	473.120	500.000	-5.4
Aroclor-1016-4	5.707	5.607	5.807	480.870	500.000	-3.8
Aroclor-1016-5	6.003	5.904	6.104	473.480	500.000	-5.3
Aroclor-1260-1	7.138	7.037	7.237	468.790	500.000	-6.2
Aroclor-1260-2	7.395	7.295	7.495	494.770	500.000	-1.0
Aroclor-1260-3	7.758	7.658	7.858	467.760	500.000	-6.4
Aroclor-1260-4	7.983	7.884	8.084	478.740	500.000	-4.3
Aroclor-1260-5	8.299	8.198	8.398	492.810	500.000	-1.4
Decachlorobiphenyl	10.091	9.991	10.191	49.340	50.000	-1.3
Tetrachloro-m-xylene	4.370	4.270	4.470	50.680	50.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG NO.: P5026

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 12/02/2024

Lab Sample No.: AR1660CCC500 Data File : PO108290.D Time Analyzed: 15:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.620	4.820	545.260	500.000	9.1
Aroclor-1016-2	4.736	4.639	4.839	551.390	500.000	10.3
Aroclor-1016-3	4.912	4.815	5.015	552.780	500.000	10.6
Aroclor-1016-4	4.953	4.856	5.056	549.950	500.000	10.0
Aroclor-1016-5	5.167	5.070	5.270	547.430	500.000	9.5
Aroclor-1260-1	6.198	6.100	6.300	522.810	500.000	4.6
Aroclor-1260-2	6.385	6.288	6.488	515.130	500.000	3.0
Aroclor-1260-3	6.538	6.441	6.641	511.790	500.000	2.4
Aroclor-1260-4	7.009	6.912	7.112	492.000	500.000	-1.6
Aroclor-1260-5	7.250	7.154	7.354	475.990	500.000	-4.8
Decachlorobiphenyl	8.630	8.535	8.735	41.760	50.000	-16.5
Tetrachloro-m-xylene	3.637	3.539	3.739	57.370	50.000	14.7