

Surrogate Summary

SDG No.: **P5026**

Client: **Tully Construction Co., Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108037.D	PIBLK-PO108037.D	Tetrachloro-m-xylene	1	20	21.2	106		60	140
		Decachlorobiphenyl	1	20	23.2	116		60	140
		Tetrachloro-m-xylene	2	20	21.0	105		60	140
		Decachlorobiphenyl	2	20	22.5	112		60	140
I.BLK-PO108279.D	PIBLK-PO108279.D	Tetrachloro-m-xylene	1	20	22.8	114		60	140
		Decachlorobiphenyl	1	20	25.1	126		60	140
		Tetrachloro-m-xylene	2	20	24.0	120		60	140
		Decachlorobiphenyl	2	20	22.1	111		60	140
PB165309BL	PB165309BL	Tetrachloro-m-xylene	1	20	21.6	108		32	144
		Decachlorobiphenyl	1	20	23.3	116		32	175
		Tetrachloro-m-xylene	2	20	22.3	111		32	144
		Decachlorobiphenyl	2	20	20.9	105		32	175
PB165309BS	PB165309BS	Tetrachloro-m-xylene	1	20	21.2	106		32	144
		Decachlorobiphenyl	1	20	22.9	115		32	175
		Tetrachloro-m-xylene	2	20	20.6	103		32	144
		Decachlorobiphenyl	2	20	20.6	103		32	175
P5026-01	SOIL-1-HAM	Tetrachloro-m-xylene	1	20	18.9	94		32	144
		Decachlorobiphenyl	1	20	15.5	78		32	175
		Tetrachloro-m-xylene	2	20	21.2	106		32	144
		Decachlorobiphenyl	2	20	12.3	62		32	175
P5026-05		Tetrachloro-m-xylene	1	20	17.4	87		32	144
		Decachlorobiphenyl	1	20	14.3	72		32	175
		Tetrachloro-m-xylene	2	20	20.6	103		32	144
		Decachlorobiphenyl	2	20	12.0	60		32	175
P5026-05MS	SOIL-1-HAMMS	Tetrachloro-m-xylene	1	20	18.6	93		32	144
		Decachlorobiphenyl	1	20	16.5	82		32	175
		Tetrachloro-m-xylene	2	20	20.0	100		32	144
		Decachlorobiphenyl	2	20	13.5	68		32	175
P5026-05MSD	SOIL-1-HAMMSD	Tetrachloro-m-xylene	1	20	19.0	95		32	144
		Decachlorobiphenyl	1	20	16.4	82		32	175
		Tetrachloro-m-xylene	2	20	19.8	99		32	144
		Decachlorobiphenyl	2	20	13.8	69		32	175
I.BLK-PO108294.D	PIBLK-PO108294.D	Tetrachloro-m-xylene	1	20	22.9	115		60	140
		Decachlorobiphenyl	1	20	24.8	124		60	140
		Tetrachloro-m-xylene	2	20	24.4	122		60	140
		Decachlorobiphenyl	2	20	20.8	104		60	140