

**TPH GC CONTINUING CALIBRATION SUMMARY**

**50 PPM TRPH STD**

Lab Name: Chemtech Contract: TULL02  
ProjectID: Hamilton  
Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG No.: P5026  
DataFile: FG014922.D Analyst Name: YP\AJ Analyst Date: 12-02-2024

| Conc. (PPM) | Area Count | RF     | Average RF | %D    |
|-------------|------------|--------|------------|-------|
| 850         | 101500676  | 119413 | 117055     | 2.014 |