

SOIL TPH GC SURROGATE RECOVERY

Lab Name: Chemtech Client: Tully Construction Co., Inc.
Lab Code: CHEM Case No.: P5026 SAS No.: P5026 SDG No.: P5026

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FG014907.D	85				0
PIBLK-FG014914.D	87				0
PIBLK-FG014921.D	90				0
SOIL-1-HAM-TPH2	49				0
SOIL-1-HAM-TPH2	71				0
SOIL-1-HAM-TPH2MS	64				0
SOIL-1-HAM-TPH2MSD	61				0
PB165318BL	82				0
PB165318BS	83				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130
For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out