

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5026

Client: Tully Construction Co., Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: P5026-05MSD		Client Sample ID: SOIL-1-HAMMSD		DataFile: BF140698.D							
Naphthalene	2100	0	2200	ug/Kg	105		15		72	110	20
Acenaphthylene	2100	0	2300	ug/Kg	110		10		79	118	20
Acenaphthene	2100	0	2200	ug/Kg	105		10		70	121	20
Fluorene	2100	0	2200	ug/Kg	105		10		68	116	20
Phenanthrene	2100	0	2300	ug/Kg	110		15		52	128	20
Anthracene	2100	0	2400	ug/Kg	114		13		62	124	20
Fluoranthene	2100	0	2400	ug/Kg	114		13		44	125	20
Pyrene	2100	0	2200	ug/Kg	105		10		26	142	20
Benzo(a)anthracene	2100	0	2300	ug/Kg	110		10		71	114	20
Chrysene	2100	0	2300	ug/Kg	110		15		57	121	20
Benzo(b)fluoranthene	2100	0	2400	ug/Kg	114		13		67	121	20
Benzo(k)fluoranthene	2100	0	2500	ug/Kg	119		28	*	57	134	20
Benzo(a)pyrene	2100	0	2700	ug/Kg	129		21	*	70	142	20
Indeno(1,2,3-cd)pyrene	2100	0	2200	ug/Kg	105		5		40	129	20
Dibenz(a,h)anthracene	2100	0	2100	ug/Kg	100		5		43	123	20
Benzo(g,h,i)perylene	2100	0	2000	ug/Kg	95		5		24	125	20