

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: **P5026**

Client: **Tully Construction Co., Inc.**

Analytical Method: **SW8260D**

Datafile : VY020482.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1202SBSD01	Dichlorodifluoromethane	20	19.2	ug/Kg	96	11		64	136	20
	Chloromethane	20	19.3	ug/Kg	97	12		70	130	20
	Vinyl chloride	20	19.6	ug/Kg	98	12		72	129	20
	Bromomethane	20	20.2	ug/Kg	101	12		58	141	20
	Chloroethane	20	20.4	ug/Kg	102	13		69	130	20
	Trichlorofluoromethane	20	20.9	ug/Kg	104	12		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	20.8	ug/Kg	104	9		81	123	20
	1,1-Dichloroethene	20	20.4	ug/Kg	102	11		79	121	20
	Acetone	100	96.8	ug/Kg	97	17		60	131	20
	Carbon disulfide	20	18.7	ug/Kg	94	15		45	154	20
	Methyl tert-butyl Ether	20	22.2	ug/Kg	111	18		77	129	20
	Methyl Acetate	20	22.5	ug/Kg	113	18		69	149	20
	Methylene Chloride	20	21.8	ug/Kg	109	18		56	174	20
	trans-1,2-Dichloroethene	20	20.9	ug/Kg	104	13		80	123	20
	1,1-Dichloroethane	20	21.2	ug/Kg	106	14		82	123	20
	Cyclohexane	20	20.1	ug/Kg	101	13		76	122	20
	2-Butanone	100	110	ug/Kg	110	24	*	69	131	20
	Carbon Tetrachloride	20	21.8	ug/Kg	109	15		76	129	20
	cis-1,2-Dichloroethene	20	21.2	ug/Kg	106	13		82	123	20
	Bromochloromethane	20	19.7	ug/Kg	99	15		80	127	20
	Chloroform	20	21.7	ug/Kg	109	17		82	125	20
	1,1,1-Trichloroethane	20	21.6	ug/Kg	108	14		80	126	20
	Methylcyclohexane	20	21.6	ug/Kg	108	15		77	123	20
	Benzene	20	21.5	ug/Kg	108	16		84	121	20
	1,2-Dichloroethane	20	22.4	ug/Kg	112	17		81	126	20
	Trichloroethene	20	21.9	ug/Kg	110	16		83	122	20
	1,2-Dichloropropane	20	22.1	ug/Kg	111	19		83	122	20
	Bromodichloromethane	20	21.9	ug/Kg	110	17		82	123	20
	4-Methyl-2-Pentanone	100	120	ug/Kg	120	26	*	70	135	20
	Toluene	20	22.1	ug/Kg	111	16		83	122	20
	t-1,3-Dichloropropene	20	21.9	ug/Kg	110	18		78	124	20
	cis-1,3-Dichloropropene	20	22.1	ug/Kg	111	19		81	122	20
	1,1,2-Trichloroethane	20	22.7	ug/Kg	114	17		82	125	20
	2-Hexanone	100	110	ug/Kg	110	23	*	66	138	20
	Dibromochloromethane	20	22.0	ug/Kg	110	15		79	125	20
	1,2-Dibromoethane	20	22.6	ug/Kg	113	20		80	125	20
	Tetrachloroethene	20	21.5	ug/Kg	108	12		83	125	20
	Chlorobenzene	20	22.2	ug/Kg	111	18		84	122	20
	Ethyl Benzene	20	22.0	ug/Kg	110	16		82	124	20
	m/p-Xylenes	40	44.2	ug/Kg	111	17		83	124	20
	o-Xylene	20	21.9	ug/Kg	110	15		83	123	20
	Styrene	20	21.9	ug/Kg	110	16		82	124	20
	Bromoform	20	22.7	ug/Kg	114	20		75	127	20
	Isopropylbenzene	20	22.1	ug/Kg	111	18		82	124	20
	1,1,2,2-Tetrachloroethane	20	22.7	ug/Kg	114	24	*	77	127	20
	1,3-Dichlorobenzene	20	22.4	ug/Kg	112	17		83	122	20
	1,4-Dichlorobenzene	20	22.6	ug/Kg	113	18		84	121	20
	1,2-Dichlorobenzene	20	22.5	ug/Kg	113	19		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5026
Client: Tully Construction Co., Inc.
Analytical Method: SW8260D **Datafile :** VY020482.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1202SBSD01	1,2-Dibromo-3-Chloropropane	20	22.7	ug/Kg	114	21	*	66	134	20
	1,2,4-Trichlorobenzene	20	22.1	ug/Kg	111	19		78	127	20
	1,2,3-Trichlorobenzene	20	21.9	ug/Kg	110	17		70	137	20