

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134114 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV003				ID: CCV019				ID: CCV020		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	95	2	400000	400000	101	0	410000	102	0
Antimony	1000	920	92	0	5000	4800	97	0	4800	97	0
Arsenic	1000	1100	109	1	5000	5000	99	0	5000	101	0
Barium	520	510	99	0	10000	10000	103	0	10000	103	0
Beryllium	510	510	100	2	500	510	102	1	510	103	0
Cadmium	510	490	97	0	2500	2500	99	0	2500	99	0
Calcium	10000	10000	100	1	400000	400000	101	0	410000	101	0
Chromium	520	530	102	0	15500	15000	99	0	15000	100	0
Cobalt	520	490	95	0	2500	2500	100	1	2500	100	0
Copper	510	490	95	1	15000	15000	101	0	15000	101	0
Iron	10000	10000	104	1	400000	410000	103	0	410000	103	0
Lead	1000	960	96	0	25000	25000	99	0	25000	98	0
Magnesium	6000	6100	102	0	400000	410000	102	0	410000	102	0
Manganese	520	530	101	0	15000	15000	101	0	15000	101	0
Nickel	530	500	94	0	2500	2500	100	0	2500	100	0
Potassium	9900	9300	94	2	150000	150000	102	0	160000	104	0
Selenium	1000	900	90	1	5000	4700	94	1	4600	93	0
Silver	250	240	96	0	1250	1300	101	1	1300	102	0
Sodium	10000	9400	94	2	400000	410000	101	0	410000	103	1
Thallium	1000	1000	100	1	5000	4900	98	0	4900	98	0
Vanadium	500	510	102	2	2500	2500	102	1	2500	101	0
Zinc	1000	970	97	0	15000	14000	96	0	14000	96	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134114 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV021				ID: CCV022		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	1	410000	103	0
Antimony					5000	4900	97	0	4900	97	0
Arsenic					5000	5000	101	0	5100	102	0
Barium					10000	10000	102	1	10000	105	0
Beryllium					500	510	102	0	530	106	1
Cadmium					2500	2500	101	0	2500	100	0
Calcium					400000	400000	100	0	410000	103	0
Chromium					15500	16000	103	3	16000	101	0
Cobalt					2500	2500	101	0	2500	101	0
Copper					15000	15000	100	1	15000	103	0
Iron					400000	400000	101	1	420000	104	0
Lead					25000	25000	100	0	25000	100	0
Magnesium					400000	400000	100	0	410000	104	1
Manganese					15000	15000	99	0	15000	102	1
Nickel					2500	2500	101	0	2500	100	0
Potassium					150000	150000	102	1	160000	107	1
Selenium					5000	4900	99	1	4900	98	0
Silver					1250	1200	100	2	1200	99	0
Sodium					400000	410000	102	1	430000	106	1
Thallium					5000	4900	98	0	4900	98	0
Vanadium					2500	2500	100	1	2600	103	1
Zinc					15000	15000	101	3	15000	99	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134114 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV023				ID: CCV024		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	1	390000	96	0
Antimony					5000	4700	93	0	4600	92	0
Arsenic					5000	4900	98	0	4900	98	0
Barium					10000	10000	101	1	9800	98	0
Beryllium					500	510	102	1	490	98	1
Cadmium					2500	2400	97	0	2400	95	0
Calcium					400000	400000	100	1	390000	98	0
Chromium					15500	15000	97	0	15000	95	0
Cobalt					2500	2400	98	0	2400	96	0
Copper					15000	15000	100	0	15000	97	0
Iron					400000	400000	100	1	390000	98	0
Lead					25000	24000	96	0	24000	94	0
Magnesium					400000	400000	100	1	390000	98	0
Manganese					15000	15000	99	1	14000	97	0
Nickel					2500	2400	97	0	2400	95	0
Potassium					150000	150000	103	0	150000	101	0
Selenium					5000	4700	95	0	4600	93	0
Silver					1250	1200	95	0	1200	93	0
Sodium					400000	410000	102	1	400000	99	1
Thallium					5000	4700	94	0	4600	92	0
Vanadium					2500	2500	99	1	2400	97	1
Zinc					15000	14000	95	1	14000	93	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134114 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV025				ID: CCV026		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	390000	97	0	390000	97	0
Antimony					5000	4600	93	0	4600	92	0
Arsenic					5000	4900	98	0	4900	97	0
Barium					10000	10000	100	0	9900	99	0
Beryllium					500	500	101	1	500	99	1
Cadmium					2500	2400	96	0	2400	96	0
Calcium					400000	390000	99	0	390000	98	0
Chromium					15500	15000	96	0	15000	96	0
Cobalt					2500	2400	97	0	2400	96	0
Copper					15000	15000	98	0	15000	98	0
Iron					400000	390000	99	1	390000	98	0
Lead					25000	24000	95	0	24000	95	0
Magnesium					400000	400000	99	1	390000	98	0
Manganese					15000	15000	97	0	14000	96	1
Nickel					2500	2400	96	0	2400	96	0
Potassium					150000	150000	102	1	150000	101	0
Selenium					5000	4700	94	0	4700	93	0
Silver					1250	1200	94	0	1200	94	0
Sodium					400000	410000	101	0	400000	100	1
Thallium					5000	4700	93	0	4600	93	0
Vanadium					2500	2500	98	0	2400	98	1
Zinc					15000	14000	95	0	14000	94	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134114 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV027				ID: CCV028		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	370000	93	0	380000	96	2
Antimony					5000	4400	89	0	4600	92	0
Arsenic					5000	4800	97	0	4900	97	0
Barium					10000	9500	95	0	9800	98	2
Beryllium					500	480	95	0	490	98	1
Cadmium					2500	2300	93	0	2400	96	0
Calcium					400000	380000	95	0	390000	97	2
Chromium					15500	14000	93	0	15000	95	0
Cobalt					2500	2300	94	0	2400	96	0
Copper					15000	14000	94	0	14000	96	2
Iron					400000	370000	93	0	380000	95	2
Lead					25000	23000	92	0	24000	95	0
Magnesium					400000	380000	94	0	380000	96	2
Manganese					15000	14000	93	0	14000	94	3
Nickel					2500	2300	93	0	2400	96	0
Potassium					150000	150000	97	0	150000	99	2
Selenium					5000	4500	91	0	4700	94	0
Silver					1250	1100	91	0	1200	93	1
Sodium					400000	380000	96	0	400000	99	2
Thallium					5000	4500	90	0	4600	92	0
Vanadium					2500	2300	93	0	2400	96	3
Zinc					15000	14000	91	0	14000	93	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134114 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV029				ID: CCV030		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	2	380000	95	1
Antimony					5000	4700	94	2	4500	89	0
Arsenic					5000	5000	99	0	5000	100	0
Barium					10000	10000	100	2	9700	97	0
Beryllium					500	500	100	1	480	96	0
Cadmium					2500	2400	97	2	2400	95	0
Calcium					400000	400000	99	2	380000	95	0
Chromium					15500	15000	97	0	15000	94	0
Cobalt					2500	2400	98	2	2400	95	0
Copper					15000	15000	98	2	14000	95	1
Iron					400000	390000	98	2	380000	94	0
Lead					25000	24000	97	2	24000	94	0
Magnesium					400000	390000	97	2	380000	94	0
Manganese					15000	14000	96	2	14000	93	0
Nickel					2500	2400	97	2	2400	95	0
Potassium					150000	150000	101	2	150000	98	0
Selenium					5000	4800	95	2	4600	92	0
Silver					1250	1200	95	0	1100	92	0
Sodium					400000	410000	102	1	390000	98	0
Thallium					5000	4700	93	2	4600	91	0
Vanadium					2500	2400	97	2	2400	95	0
Zinc					15000	14000	96	0	14000	92	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB134114 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV031				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	380000	95	0			
Antimony					5000	4400	89	0			
Arsenic					5000	5000	100	0			
Barium					10000	9700	97	0			
Beryllium					500	480	96	1			
Cadmium					2500	2400	95	0			
Calcium					400000	380000	95	0			
Chromium					15500	15000	94	0			
Cobalt					2500	2400	95	0			
Copper					15000	14000	95	0			
Iron					400000	370000	93	0			
Lead					25000	24000	94	0			
Magnesium					400000	370000	94	0			
Manganese					15000	14000	92	0			
Nickel					2500	2400	95	0			
Potassium					150000	150000	98	0			
Selenium					5000	4600	92	0			
Silver					1250	1100	91	0			
Sodium					400000	390000	97	1			
Thallium					5000	4600	91	0			
Vanadium					2500	2400	94	1			
Zinc					15000	14000	91	1			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134132 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV005				ID: CCV049				ID: CCV050		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2500	101	1	400000	400000	101	0	410000	101	0
Antimony	1000	950	95	0	5000	4900	98	0	4800	96	0
Arsenic	1000	1100	108	1	5000	5000	99	0	4900	99	0
Barium	520	540	104	2	10000	10000	101	0	10000	101	0
Beryllium	510	540	106	2	500	510	101	1	510	101	0
Cadmium	510	510	101	0	2500	2500	100	0	2500	98	0
Calcium	10000	11000	106	2	400000	410000	103	0	410000	103	0
Chromium	520	560	108	1	15500	16000	100	0	16000	100	0
Cobalt	520	510	99	0	2500	2500	99	0	2500	98	0
Copper	510	530	105	1	15000	15000	101	0	15000	101	0
Iron	10000	11000	109	2	400000	410000	102	0	410000	103	0
Lead	1000	1000	100	0	25000	25000	99	0	24000	98	0
Magnesium	6000	6400	107	1	400000	400000	101	0	410000	102	0
Manganese	520	550	106	2	15000	15000	100	0	15000	101	0
Nickel	530	520	98	0	2500	2500	100	0	2500	99	0
Potassium	9900	9800	99	1	150000	150000	102	0	150000	102	0
Selenium	1000	1000	100	1	5000	4900	99	0	4900	97	0
Silver	250	250	100	1	1250	1300	101	0	1300	102	1
Sodium	10000	9400	94	1	400000	410000	102	0	410000	103	0
Thallium	1000	1000	103	0	5000	4800	96	0	4800	95	0
Vanadium	500	540	107	5	2500	2500	99	1	2500	100	0
Zinc	1000	980	98	2	15000	15000	99	1	15000	99	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB134132 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV051				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	1			
Antimony					5000	4900	98	0			
Arsenic					5000	5000	99	0			
Barium					10000	10000	101	1			
Beryllium					500	500	101	2			
Cadmium					2500	2500	99	0			
Calcium					400000	410000	102	1			
Chromium					15500	16000	100	0			
Cobalt					2500	2500	99	0			
Copper					15000	15000	100	1			
Iron					400000	410000	102	1			
Lead					25000	25000	98	0			
Magnesium					400000	400000	101	1			
Manganese					15000	15000	100	1			
Nickel					2500	2500	100	0			
Potassium					150000	150000	102	1			
Selenium					5000	4900	98	0			
Silver					1250	1300	101	0			
Sodium					400000	410000	103	2			
Thallium					5000	4800	96	0			
Vanadium					2500	2500	99	0			
Zinc					15000	15000	99	1			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134139 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV006				ID: CCV053				ID: CCV054		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2500	99	1	400000	400000	100	0	400000	100	0
Antimony	1000	950	95	0	5000	4900	98	0	5000	99	0
Arsenic	1000	1100	109	1	5000	4900	98	1	4900	99	1
Barium	520	530	102	0	10000	10000	100	0	10000	101	0
Beryllium	510	540	107	1	500	510	101	1	510	102	0
Cadmium	510	520	103	0	2500	2500	98	0	2500	100	0
Calcium	10000	11000	106	0	400000	400000	101	0	410000	102	0
Chromium	520	570	109	0	15500	15000	99	0	16000	101	0
Cobalt	520	520	100	0	2500	2500	99	0	2500	101	0
Copper	510	520	102	1	15000	15000	99	0	15000	100	0
Iron	10000	11000	107	1	400000	410000	102	0	410000	102	0
Lead	1000	1000	101	0	25000	24000	98	0	25000	99	0
Magnesium	6000	6300	105	2	400000	400000	101	0	400000	100	0
Manganese	520	540	103	1	15000	15000	99	0	15000	99	0
Nickel	530	530	100	0	2500	2500	99	0	2500	101	0
Potassium	9900	9800	99	2	150000	150000	101	0	150000	102	0
Selenium	1000	1000	100	1	5000	4900	98	0	5000	99	0
Silver	250	250	100	0	1250	1300	100	0	1300	101	0
Sodium	10000	9300	93	1	400000	410000	104	0	410000	103	0
Thallium	1000	1000	103	0	5000	4800	96	0	4900	97	0
Vanadium	500	530	107	2	2500	2500	99	0	2500	99	1
Zinc	1000	1000	101	0	15000	15000	100	0	15000	100	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJJ2Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134151 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV007				ID: CCV055				ID: CCV056		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	95	3	400000	400000	100	0	420000	104	1
Antimony	1000	1000	103	0	5000	4900	98	0	5100	103	1
Arsenic	1000	1100	107	0	5000	4900	99	0	5100	102	0
Barium	520	530	102	0	10000	10000	102	1	10000	104	1
Beryllium	510	540	107	1	500	520	104	1	530	105	1
Cadmium	510	510	100	0	2500	2500	99	0	2600	103	0
Calcium	10000	10000	104	1	400000	410000	103	0	420000	106	1
Chromium	520	570	109	0	15500	16000	101	0	16000	103	0
Cobalt	520	510	98	0	2500	2500	100	0	2600	104	0
Copper	510	530	104	0	15000	15000	102	0	16000	104	1
Iron	10000	11000	107	1	400000	410000	103	0	430000	108	0
Lead	1000	990	99	0	25000	25000	99	0	26000	102	0
Magnesium	6000	6200	104	2	400000	410000	102	1	420000	106	1
Manganese	520	550	105	1	15000	15000	102	0	16000	106	1
Nickel	530	520	98	0	2500	2500	100	0	2600	104	0
Potassium	9900	11000	109	1	150000	160000	104	1	160000	107	1
Selenium	1000	1000	100	1	5000	5000	99	0	5200	103	1
Silver	250	250	102	1	1250	1300	102	1	1300	103	0
Sodium	10000	10000	101	0	400000	420000	104	1	430000	106	1
Thallium	1000	1000	102	0	5000	4800	96	0	5000	100	0
Vanadium	500	460	92	0	2500	2500	102	0	2600	105	0
Zinc	1000	1000	101	1	15000	15000	101	0	16000	107	0