

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501		
ICV Standard	MP83508		
CCV Standard	MP83511		
ICSA Standard	MP83509,MP83510		
CRI Standard			
LCS Standard			
Chk Standard	MP83513,MP83514		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/10/24 15:54		Kareem	OK
2	S1	S01	CAL2	12/10/24 15:58		Kareem	OK
3	S2	S02	CAL3	12/10/24 16:02		Kareem	OK
4	S3	S03	CAL4	12/10/24 16:06		Kareem	OK
5	S4	S04	CAL5	12/10/24 16:09		Kareem	OK
6	S5	S05	CAL6	12/10/24 16:13		Kareem	OK
7	S6	S06	CAL7	12/10/24 16:17		Kareem	OK
8	ICV014	ICV014	ICV	12/10/24 16:45		Kareem	OK
9	ICB014	ICB014	ICB	12/10/24 16:51		Kareem	OK
10	ICSA014	ICSA014	ICSA	12/10/24 17:21		Kareem	OK
11	ICSAB014	ICSAB014	ICSAB	12/10/24 17:27		Kareem	OK
12	CCV089	CCV089	CCV	12/10/24 17:48		Kareem	OK
13	CCB089	CCB089	CCB	12/10/24 17:52		Kareem	OK
14	PB165364BL	PBS364	MB	12/10/24 17:56		Kareem	OK
15	PB165351BL	PBS351	MB	12/10/24 18:04		Kareem	OK
16	PB165364BS	LCS364	LCS	12/10/24 18:09		Kareem	OK
17	PB165351BS	LCS351	LCS	12/10/24 18:13		Kareem	OK
18	PB165350BL	PBS350	MB	12/10/24 18:17		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	
Chk Standard	MP83513,MP83514

19	PB165350BS	LCS350	LCS	12/10/24 18:22		Kareem	OK
20	CCV090	CCV090	CCV	12/10/24 18:26		Kareem	OK
21	CCB090	CCB090	CCB	12/10/24 18:29		Kareem	OK
22	P5041-01	MBHK78	SAM	12/10/24 18:33		Kareem	OK
23	P5041-02	MBHK78D	DUP	12/10/24 18:37		Kareem	OK
24	P5041-01L	MBHK78L	SD	12/10/24 18:41		Kareem	OK
25	P5041-03	MBHK78S	MS	12/10/24 18:45	MS Fail For Se(Below RL)	Kareem	OK
26	P5042-01	MBHKB0	SAM	12/10/24 18:49		Kareem	OK
27	P5042-02	MBHKB0D	DUP	12/10/24 18:53		Kareem	OK
28	P5042-01L	MBHKB0L	SD	12/10/24 18:56		Kareem	OK
29	P5042-03	MBHKB0S	MS	12/10/24 19:00	MS Fail For Se(Below RL)	Kareem	OK
30	P5043-01	MBHKM2	SAM	12/10/24 19:04		Kareem	OK
31	P5043-02	MBHKM2D	DUP	12/10/24 19:08		Kareem	OK
32	P5043-01L	MBHKM2L	SD	12/10/24 19:12		Kareem	OK
33	P5043-03	MBHKM2S	MS	12/10/24 19:16	MS Fail For Ag	Kareem	OK
34	P4977-03	ME29F8	SAM	12/10/24 19:20	Ca High	Kareem	Dilution
35	P4977-04	ME29F8D	DUP	12/10/24 19:23	Ca High	Kareem	Dilution
36	P4977-03L	ME29F8L	SD	12/10/24 19:27	Ca High	Kareem	Dilution
37	P4977-05	ME29F8S	MS	12/10/24 19:31		Kareem	OK
38	P4976-09	ME29E4D	DUP	12/10/24 19:35	Ca High	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501		
ICV Standard	MP83508		
CCV Standard	MP83511		
ICSA Standard	MP83509,MP83510		
CRI Standard			
LCS Standard			
Chk Standard	MP83513,MP83514		

39	P4976-08L	ME29E4L	SD	12/10/24 19:39	Ca High	Kareem	Dilution
40	P4976-10	ME29E4S	MS	12/10/24 19:43		Kareem	OK
41	P5040-01	MBHKD6	SAM	12/10/24 19:47		Kareem	OK
42	CCV091	CCV091	CCV	12/10/24 19:51		Kareem	OK
43	CCB091	CCB091	CCB	12/10/24 19:54		Kareem	OK
44	P5040-02	MBHKD6D	DUP	12/10/24 19:58		Kareem	OK
45	P5040-01L	MBHKD6L	SD	12/10/24 20:02		Kareem	OK
46	P5040-03	MBHKD6S	MS	12/10/24 20:06	MS Fail For Se(Below RL)	Kareem	OK
47	P4977-01	ME29F6	SAM	12/10/24 20:10	Ca High	Kareem	Dilution
48	P4977-02	ME29F7	SAM	12/10/24 20:14	Ca High	Kareem	Dilution
49	P4977-06	ME29F9	SAM	12/10/24 20:18	Ca High	Kareem	Dilution
50	P4977-07	ME29G0	SAM	12/10/24 20:21	Ca High	Kareem	Dilution
51	P4977-08	ME29G1	SAM	12/10/24 20:25	Ca High	Kareem	Dilution
52	P4977-09	ME29G2	SAM	12/10/24 20:29	Ca High	Kareem	Dilution
53	P4976-01	ME29D7	SAM	12/10/24 20:33	Ca High	Kareem	Dilution
54	P4976-02	ME29D8	SAM	12/10/24 20:37	Ca High	Kareem	Dilution
55	P4976-03	ME29D9	SAM	12/10/24 20:41	Ca High	Kareem	Dilution
56	P4976-04	ME29E0	SAM	12/10/24 20:45		Kareem	OK
57	P4976-05	ME29E1	SAM	12/10/24 20:48	Ca High	Kareem	Dilution
58	P4976-06	ME29E2	SAM	12/10/24 20:52	Ca High	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	
Chk Standard	MP83513,MP83514

59	P4976-07	ME29E3	SAM	12/10/24 20:56	Ca High	Kareem	Dilution
60	P4976-08	ME29E4	SAM	12/10/24 21:00	Ca High	Kareem	Dilution
61	P4976-11	ME29E5	SAM	12/10/24 21:04	Ca High	Kareem	Dilution
62	P4976-12	ME29E6	SAM	12/10/24 21:08	Ca High	Kareem	Dilution
63	P4976-13	ME29E7	SAM	12/10/24 21:12	Ca High	Kareem	Dilution
64	CCV092	CCV092	CCV	12/10/24 21:29		Kareem	OK
65	CCB092	CCB092	CCB	12/10/24 21:52		Kareem	OK
66	P4976-14	ME29E8	SAM	12/10/24 21:56	Ca High	Kareem	Dilution
67	P4976-15	ME29E9	SAM	12/10/24 22:00	Ca High	Kareem	Dilution
68	P4976-16	ME29F0	SAM	12/10/24 22:04	Ca High	Kareem	Dilution
69	P4976-17	ME29F1	SAM	12/10/24 22:08	Ca High	Kareem	Dilution
70	P4976-18	ME29F2	SAM	12/10/24 22:12	Ca High	Kareem	Dilution
71	P4976-19	ME29F3	SAM	12/10/24 22:16	Ca High	Kareem	Dilution
72	P4976-20	ME29F4	SAM	12/10/24 22:20	Ca High	Kareem	Dilution
73	P4976-21	ME29F5	SAM	12/10/24 22:23	Ca High	Kareem	Dilution
74	P4880-14	MBHH91	SAM	12/10/24 22:27		Kareem	OK
75	P4880-15	MBHH91D	DUP	12/10/24 22:31		Kareem	OK
76	P4880-14L	MBHH91L	SD	12/10/24 22:35		Kareem	OK
77	P4880-16	MBHH91S	MS	12/10/24 22:39	MS Fail For Ag,Se(Below RL)	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501		
ICV Standard	MP83508		
CCV Standard	MP83511		
ICSA Standard	MP83509,MP83510		
CRI Standard			
LCS Standard			
Chk Standard	MP83513,MP83514		

78	P4880-12DL	MBHH89	SAM	12/10/24 22:43	2X For Ca	Kareem	Confirms
79	P4905-02	MBHHA6	SAM	12/10/24 22:46		Kareem	OK
80	P4905-03	MBHHA6D	DUP	12/10/24 22:50		Kareem	OK
81	P4905-02L	MBHHA6L	SD	12/10/24 22:54		Kareem	OK
82	P4905-04	MBHHA6S	MS	12/10/24 22:58	MS Fail For Sb,As(115.87ppb),Ba(2331 .33ppb),Cu(457.90),Se(Bel ow RL),Ag,Zn(2281.14ppb)	Kareem	OK
83	CCV093	CCV093	CCV	12/10/24 23:02		Kareem	OK
84	CCB093	CCB093	CCB	12/10/24 23:05		Kareem	OK
85	P4907-01A	MBHHE1A	PS	12/10/24 23:09	Not Use	Kareem	Not Ok
86	P4912-04A	MBHHJ4A	PS	12/10/24 23:13	PS For Cu,Zn,Se	Kareem	OK
87	P4912-14	MBHHK2	SAM	12/10/24 23:17		Kareem	OK
88	P4912-15	MBHHK3	SAM	12/10/24 23:21		Kareem	OK
89	P4912-16	MBHHL8	SAM	12/10/24 23:24		Kareem	OK
90	P4912-17	MBHHL9	SAM	12/10/24 23:28		Kareem	OK
91	P4912-18	MBHHM0	SAM	12/10/24 23:32		Kareem	OK
92	P4912-19	MBHHM1	SAM	12/10/24 23:36		Kareem	OK
93	P4912-20	MBHHM2	SAM	12/10/24 23:40		Kareem	OK
94	P4912-21	MBHHM3	SAM	12/10/24 23:43		Kareem	OK
95	P4912-22	MBHHM4	SAM	12/10/24 23:47		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	
Chk Standard	MP83513,MP83514

96	P4914-11A	MBHHK4A	PS	12/10/24 23:51	PS For Cu	Kareem	OK
97	P4915-08A	MBHHM8A	PS	12/10/24 23:55	PS For Se - 0.07ml mp83567-10ml sample	Kareem	OK
98	P4915-01	MBHHL4	SAM	12/10/24 23:58		Kareem	OK
99	P4915-02	MBHHL5	SAM	12/11/24 00:02		Kareem	OK
100	P4915-03	MBHHL6	SAM	12/11/24 00:06		Kareem	OK
101	P4915-04	MBHHL7	SAM	12/11/24 00:10		Kareem	OK
102	P4915-05	MBHHM5	SAM	12/11/24 00:14		Kareem	OK
103	P4915-06	MBHHM6	SAM	12/11/24 00:18	Not Use	Kareem	Not Ok
104	P4932-02	MBHHQ2	SAM	12/11/24 00:22	Not Use	Kareem	Not Ok
105	CCV094	CCV094	CCV	12/11/24 00:25		Kareem	OK
106	CCB094	CCB094	CCB	12/11/24 00:29		Kareem	OK
107	P4932-01	MBHHQ1	SAM	12/11/24 00:33		Kareem	OK
108	P4932-05	MBHHQ3	SAM	12/11/24 00:37		Kareem	OK
109	P4932-06	MBHHQ4	SAM	12/11/24 00:41		Kareem	OK
110	P4932-07	MBHHQ5	SAM	12/11/24 00:44		Kareem	OK
111	P4932-08	MBHHQ6	SAM	12/11/24 00:48		Kareem	OK
112	P4932-09	MBHHQ7	SAM	12/11/24 00:52		Kareem	OK
113	P4932-10	MBHHQ8	SAM	12/11/24 00:56		Kareem	OK
114	P4932-11	MBHHQ9	SAM	12/11/24 00:59		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	
Chk Standard	MP83513,MP83514

115	P4932-12	MBHHP1	SAM	12/11/24 01:03		Kareem	OK
116	P4932-13	MBHHP2	SAM	12/11/24 01:07		Kareem	OK
117	P4932-14	MBHHP3	SAM	12/11/24 01:11		Kareem	OK
118	P4932-15	MBHHP4	SAM	12/11/24 01:15		Kareem	OK
119	P4932-16	MBHHP5	SAM	12/11/24 01:18		Kareem	OK
120	P4932-17	MBHHP6	SAM	12/11/24 01:22		Kareem	OK
121	P4932-18	MBHHP7	SAM	12/11/24 01:26		Kareem	OK
122	P4932-19	MBHHP8	SAM	12/11/24 01:30		Kareem	OK
123	P4932-20	MBHHP9	SAM	12/11/24 01:34		Kareem	OK
124	P4932-21	MBHHQ0	SAM	12/11/24 01:37		Kareem	OK
125	P4932-22	MBHHR0	SAM	12/11/24 01:41		Kareem	OK
126	P4933-05DL	MBHHR5	SAM	12/11/24 01:45	5X For Mn	Kareem	Confirms
127	CCV095	CCV095	CCV	12/11/24 01:49		Kareem	OK
128	CCB095	CCB095	CCB	12/11/24 01:52		Kareem	OK
129	P4933-18	MBHHX2	SAM	12/11/24 01:56		Kareem	OK
130	P4933-19	MBHHX3	SAM	12/11/24 02:00		Kareem	OK
131	P4933-20	MBHHX4	SAM	12/11/24 02:04		Kareem	OK
132	P4933-21	MBHHX4D	DUP	12/11/24 02:08		Kareem	OK
133	P4933-20L	MBHHX4L	SD	12/11/24 02:12		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133883

Review By	Sarabjit Jaswal	Review On	12/12/2024 8:19:04 AM
Supervise By	Mohan Bera	Supervise On	12/12/2024 8:19:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	
Chk Standard	MP83513,MP83514

134	P4933-22	MBHHX4S	MS	12/11/24 02:15	MS Fail for Se(Below RL),Zn(1944.24ppb)	Kareem	OK
135	P4939-10A	MBHHW8A	PS	12/11/24 02:19	PS For Se	Kareem	OK
136	P4939-22DL	MBHHZ2	SAM	12/11/24 02:23	2X For Cu	Kareem	Confirms
137	P5007-15DL	MBHK00	SAM	12/11/24 02:27	2X For Ca	Kareem	Confirms
138	P5007-10A	MBHJZ7A	PS	12/11/24 02:31	PS For As	Kareem	OK
139	P5014-12A	MBHJY0A	PS	12/11/24 02:34	PS For As,Se	Kareem	OK
140	P5015-17A	MBHKQ6A	PS	12/11/24 02:38	PS For As,Se,Tl,Zn	Kareem	OK
141	P5015-05DL	MBHKK0	SAM	12/11/24 02:42	2X For Pb	Kareem	Confirms
142	P5015-08DL	MBHKK3	SAM	12/11/24 02:46	2X For Ca	Kareem	Confirms
143	P5015-09DL	MBHKK4	SAM	12/11/24 02:50	2X For Ca	Kareem	Confirms
144	P5015-14DL	MBHKQ2	SAM	12/11/24 02:54	2X For Ca	Kareem	Confirms
145	CCV096	CCV096	CCV	12/11/24 02:57		Kareem	OK
146	CCB096	CCB096	CCB	12/11/24 03:01		Kareem	OK