

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** P5044
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** P5044 **SAS No.:** P5044
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	512	510	100	95 - 105	P	12/05/2024	13:53	LB133769
	Zinc	1020	1000	102	95 - 105	P	12/05/2024	13:53	LB133769

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** P5044
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** P5044 **SAS No.:** P5044
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.8	20.0	109	80 - 120	P	12/05/2024	14:05	LB133769
	Zinc	42.9	40.0	107	80 - 120	P	12/05/2024	14:05	LB133769

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** P5044
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** P5044 **SAS No.:** P5044
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	12/05/2024	14:45	LB133769
	Zinc	2510	2500	100	90 - 110	P	12/05/2024	14:45	LB133769
CCV02	Copper	1200	1250	96	90 - 110	P	12/05/2024	15:34	LB133769
	Zinc	2450	2500	98	90 - 110	P	12/05/2024	15:34	LB133769
CCV03	Copper	1220	1250	98	90 - 110	P	12/05/2024	16:42	LB133769
	Zinc	2520	2500	101	90 - 110	P	12/05/2024	16:42	LB133769
CCV04	Copper	1170	1250	94	90 - 110	P	12/05/2024	17:41	LB133769
	Zinc	2380	2500	95	90 - 110	P	12/05/2024	17:41	LB133769
CCV05	Copper	1140	1250	91	90 - 110	P	12/05/2024	18:39	LB133769
	Zinc	2390	2500	96	90 - 110	P	12/05/2024	18:39	LB133769
CCV06	Copper	1160	1250	92	90 - 110	P	12/05/2024	19:45	LB133769
	Zinc	2360	2500	95	90 - 110	P	12/05/2024	19:45	LB133769
CCV07	Copper	1150	1250	92	90 - 110	P	12/05/2024	20:23	LB133769
	Zinc	2430	2500	97	90 - 110	P	12/05/2024	20:23	LB133769