



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: P5044

Contract: TRIS02

Lab Code: CHEM

Case No.: P5044

SAS No.: P5044

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	14:10	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	14:10	LB133769

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: P5044

Contract: TRIS02

Lab Code: CHEM

Case No.: P5044

SAS No.: P5044

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	14:50	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	14:50	LB133769
CCB02	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	15:41	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	15:41	LB133769
CCB03	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	16:46	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	16:46	LB133769
CCB04	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	17:47	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	17:47	LB133769
CCB05	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	18:43	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	18:43	LB133769
CCB06	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	19:50	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	19:50	LB133769
CCB07	Copper	20.0	+/-20.0	U	20.0	P	12/05/2024	20:27	LB133769
	Zinc	40.0	+/-40.0	U	40.0	P	12/05/2024	20:27	LB133769

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Tris Pharma, Inc.</u>				SDG No.:	<u>P5044</u>			
Contract:	<u>TRIS02</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>P5044</u>	SAS No.:	<u>P5044</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number