

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tris Pharma, Inc. **SDG No.:** P5044
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** P5044 **SAS No.:** P5044
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	4.69	2.0	234	-18	22	12/05/2024	14:18	LB133769
	Zinc	8.24			-40	40	12/05/2024	14:18	LB133769
ICSAB01	Copper	462	511	90	434	588	12/05/2024	14:33	LB133769
	Zinc	953	952	100	809	1095	12/05/2024	14:33	LB133769
ICSA	Copper	-1.04	2.0	52	-18	22	12/05/2024	14:37	LB133769
	Zinc	9.03			-40	40	12/05/2024	14:37	LB133769
ICSAB	Copper	518	511	101	434	588	12/05/2024	14:41	LB133769
	Zinc	1090	952	114	809	1095	12/05/2024	14:41	LB133769