

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TRIS02

Lab Code: CHEM Case No.: P5044 SAS No.: P5044 SDG No.: P5044

Instrument ID: MSVOA_N Calibration Date/Time: 12/06/2024 10:26

Lab File ID: VN085118.D Init. Calib. Date(s): 12/06/2024 12/06/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:02 11:37

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Ethyl Acetate	0.442	0.404		-8.6	
Isopropyl Acetate	0.768	0.746		-2.87	
Acetone	0.243	0.228		-6.17	
Methylene Chloride	2.073	2.012		-2.94	
1,2-Dichloroethane-d4	2.380	2.393		0.55	
Toluene-d8	1.407	1.443		2.56	
4-Bromofluorobenzene	0.496	0.495		-0.2	
n-amyl Acetate	0.646	0.619		-4.18	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.