

Report of Analysis

Client:	Tris Pharma, Inc.		Date Collected:	
Project:	Quarterly		Date Received:	
Client Sample ID:	VN1206WBL01		SDG No.:	P5044
Lab Sample ID:	VN1206WBL01		Matrix:	Water
Analytical Method:	E624.1		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085126.D	1		12/06/24 14:40	VN120624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
141-78-6	Ethyl Acetate	1.50	U	1.50	5.00	ug/L
108-21-4	Isopropyl Acetate	0.66	U	0.66	5.00	ug/L
67-64-1	Acetone	4.90	U	4.90	25.0	ug/L
75-09-2	Methylene Chloride	1.20	U	1.20	5.00	ug/L
628-63-7	n-amyl Acetate	0.91	U	0.91	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.7		91 - 110	102%	SPK: 30
2037-26-5	Toluene-d8	29.0		91 - 112	97%	SPK: 30
460-00-4	4-Bromofluorobenzene	24.3		63 - 112	81%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	29600	7.812			
540-36-3	1,4-Difluorobenzene	150000	9.1			
3114-55-4	Chlorobenzene-d5	136000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products