

### Surrogate Summary

SDG No.: P5044

Client: Tris Pharma, Inc.

Analytical Method: SW624.1

| Lab Sample ID | Client ID       | Parameter             | Spike | Result | RecoveryQual | Limits |      |
|---------------|-----------------|-----------------------|-------|--------|--------------|--------|------|
|               |                 |                       |       |        |              | Low    | High |
| P5044-01      | OUTFALL-DSN-001 | 1,2-Dichloroethane-d4 | 30    | 30.2   | 101          | 91     | 110  |
|               |                 | Toluene-d8            | 30    | 28.3   | 94           | 91     | 112  |
|               |                 | 4-Bromofluorobenzene  | 30    | 24.6   | 82           | 63     | 112  |
| P5044-02      | OUTFALL-DSN-002 | 1,2-Dichloroethane-d4 | 30    | 31.6   | 105          | 91     | 110  |
|               |                 | Toluene-d8            | 30    | 28.3   | 94           | 91     | 112  |
|               |                 | 4-Bromofluorobenzene  | 30    | 24.8   | 83           | 63     | 112  |
| VN1206WBL01   | VN1206WBL01     | 1,2-Dichloroethane-d4 | 30    | 30.7   | 102          | 91     | 110  |
|               |                 | Toluene-d8            | 30    | 29.0   | 97           | 91     | 112  |
|               |                 | 4-Bromofluorobenzene  | 30    | 24.3   | 81           | 63     | 112  |
| VN1206WBS01   | VN1206WBS01     | 1,2-Dichloroethane-d4 | 30    | 30.1   | 100          | 91     | 110  |
|               |                 | Toluene-d8            | 30    | 29.7   | 99           | 91     | 112  |
|               |                 | 4-Bromofluorobenzene  | 30    | 30.0   | 100          | 63     | 112  |
| VN1206WBSD0   | VN1206WBSD01    | 1,2-Dichloroethane-d4 | 30    | 30.8   | 103          | 91     | 110  |
|               |                 | Toluene-d8            | 30    | 29.6   | 99           | 91     | 112  |
|               |                 | 4-Bromofluorobenzene  | 30    | 29.6   | 99           | 63     | 112  |