

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5050 SAS No.: P5050 SDG NO.: P5050

Continuing Calib Date: 12/03/2024 Initial Calibration Date(s): 11/21/2024 11/22/2024

Continuing Calib Time: 11:06 Initial Calibration Time(s): 16:13 01:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.86	4.86	4.79	4.93	0.00
Aroclor-1016-2	(2)	4.88	4.88	4.81	4.95	0.00
Aroclor-1016-3	(3)	4.95	4.94	4.87	5.01	-0.01
Aroclor-1016-4	(4)	5.05	5.05	4.98	5.12	0.00
Aroclor-1016-5	(5)	5.36	5.36	5.29	5.43	0.00
Aroclor-1260-1	(1)	6.58	6.57	6.50	6.64	0.00
Aroclor-1260-2	(2)	6.86	6.86	6.79	6.93	0.00
Aroclor-1260-3	(3)	7.25	7.25	7.18	7.32	0.00
Aroclor-1260-4	(4)	7.49	7.49	7.42	7.56	0.00
Aroclor-1260-5	(5)	7.83	7.83	7.76	7.90	0.00
TCX		3.64	3.64	3.59	3.69	0.00
DCB		9.53	9.53	9.43	9.63	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P5050 SAS No.: P5050 SDG NO.: P5050

Continuing Calib Date: 12/03/2024 Initial Calibration Date(s): 11/21/2024 11/22/2024

Continuing Calib Time: 11:06 Initial Calibration Time(s): 16:13 01:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.09	4.09	4.02	4.16	0.00
Aroclor-1016-2	(2)	4.11	4.11	4.04	4.18	0.00
Aroclor-1016-3	(3)	4.29	4.29	4.22	4.36	0.00
Aroclor-1016-4	(4)	4.34	4.34	4.27	4.41	0.00
Aroclor-1016-5	(5)	4.56	4.56	4.49	4.63	0.00
Aroclor-1260-1	(1)	5.66	5.66	5.59	5.73	0.00
Aroclor-1260-2	(2)	5.86	5.86	5.79	5.93	0.00
Aroclor-1260-3	(3)	6.02	6.02	5.95	6.09	0.00
Aroclor-1260-4	(4)	6.53	6.53	6.46	6.60	0.00
Aroclor-1260-5	(5)	6.79	6.80	6.73	6.87	0.01
TCX		2.97	2.97	2.92	3.02	0.00
DCB		8.25	8.25	8.15	8.35	0.00