

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5050 SAS No.: P5050 SDG NO.: P5050

EPA Sample No.: SSTD020349 Date Analyzed: 12/03/2024

Lab File ID: BN035400.D Time Analyzed: 09:30

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	77198	7.293	325386	10.05	257861	13.96
UPPER LIMIT	154396	7.793	650772	10.546	515722	14.463
LOWER LIMIT	38599	6.793	162693	9.546	128931	13.463
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	78825	7.30	324440	10.05	258462	13.96

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.