

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5050 SAS No.: P5050 SDG NO.: P5050

EPA Sample No.: SSTD020349 Date Analyzed: 12/03/2024

Lab File ID: BN035400.D Time Analyzed: 09:30

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	590832	16.728	623287	20.986	753050	23.657
UPPER LIMIT	1181660	17.228	1246570	21.486	1506100	24.157
LOWER LIMIT	295416	16.228	311644	20.486	376525	23.157
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	612241	16.73	652912	20.98	752287	23.65

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.