



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P5051

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133695

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.499	0.5	100	95-105	12/02/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	12/02/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	12/02/2024

Initial and Continuing Calibration Verification

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P5051

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133772

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.097	0.099	98	85-115	12/05/2024
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	12/05/2024
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	12/05/2024
Sample ID: Cyanide	CCV3	mg/L	0.25	0.25	100	90-110	12/05/2024

Initial and Continuing Calibration Verification

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P5051

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB134007

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.01	7	100	90-110	12/02/2024
Sample ID:	CCV1						
Field pH		pH	7.01	7	100	90-110	12/02/2024
Sample ID:	CCV2						
Field pH		pH	7	7	100	90-110	12/02/2024