

Initial and Continuing Calibration Blank Summary

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P5051

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133695

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0021	0.01	12/02/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0021	0.01	12/02/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0021	0.01	12/02/2024

Initial and Continuing Calibration Blank Summary

Client: New York City DEP of Environmental Protection/BWS

SDG No.: P5051

Project: Industrial Wastewater Discharge Permit - Fall 2024

RunNo.: LB133772

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00093	0.005	12/05/2024