

Laboratory Control Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P5051
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Run No.:	LB133695

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133695BS							
Hexavalent Chromium	mg/L	0.5	0.51		102	1	90-111	12/02/2024

Laboratory Control Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P5051
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Run No.:	LB133715

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133715BS							
TSS	mg/L	550	536		98	1	90-110	12/03/2024

Laboratory Control Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P5051
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Run No.:	LB133806

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133806BS							
Non-Polar Material	mg/L	20.0	16.6		83	1	78-114	12/07/2024

Laboratory Control Sample Summary

Client:	New York City DEP of Environmental Protection/BWS	SDG No.:	P5051
Project:	Industrial Wastewater Discharge Permit - Fall 2024	Run No.:	LB133772

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165397BS							
Cyanide	mg/L	0.1	0.095		95	1	85-115	12/05/2024