


LB133852

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/1)
0.0	0.000	-0.020	-0.020	36	0.000	36					1
0.05	0.050	0.051	0.001	297	0.0 %	297	0				2
0.2	0.200	0.183	-0.017	777	0.0 %	777					9
2.5	2.500	2.383	-0.117	8781	0.0 %	8781					5
5.0	5.000	5.121	0.121	18746	0.0 %	18746					2
7.5	7.500	7.855	0.355	28696	0.0 %	28696					5
10.0	10.000	9.677	-0.323	35324	0.0 %	35324					3

LB133852 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Dec 2024 09:29:03

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	38	-	0.000	245.1	PPB	10 Dec 2024 09:37:35	S	Std
0.05 - 1	50.05	297	-	0.050	245.1	PPB	10 Dec 2024 09:42:31	S	Std
0.2 - 1	50.2	777	-	0.200	245.1	PPB	10 Dec 2024 09:44:48	S	Std
2.5 - 1	52.5	8781	-	2.500	245.1	PPB	10 Dec 2024 09:47:04	S	Std
5.0 - 1	55	18746	-	5.000	245.1	PPB	10 Dec 2024 09:49:20	S	Std
7.5 - 1	57.5	28696	-	7.500	245.1	PPB	10 Dec 2024 09:51:37	S	Std
10.0 - 1	510	35324	-	10.000	245.1	PPB	10 Dec 2024 09:59:28	S	Std
ICV14 - 1	ICV14	14778	4.0307	-245.1		PPB	10 Dec 2024 10:09:31	U	SMPL
ICB14 - 1	ICB14	-6	-0.0319	-245.1		PPB	10 Dec 2024 10:17:09	U	SMPL
CCV10 - 1	CCV10	19970	5.4574	-245.1		PPB	10 Dec 2024 10:29:43	U	SMPL
CCB10 - 1	CCB10	48	-0.0171	-245.1		PPB	10 Dec 2024 10:34:42	U	SMPL
CRA - 1	CRA	248	0.0379	-245.1		PPB	10 Dec 2024 10:39:32	U	SMPL
HighStd - 1	HighStd	31722	8.6868	-245.1		PPB	10 Dec 2024 10:49:15	U	SMPL
ChkStd - 1	ChkStd	23934	6.5467	-245.1		PPB	10 Dec 2024 10:51:31	U	SMPL
PB165519BL - 1	PBW	-208	-0.0874	-245.1		PPB	10 Dec 2024 10:53:49	U	SMPL
PB165519BS - 1	LCSW	14515	3.9584	-245.1		PPB	10 Dec 2024 11:02:20	U	SMPL
P4368-10 - 1	LOD-MDL-WATER-04-QT4-2024	233	0.0338	-245.1		PPB	10 Dec 2024 11:12:18	U	SMPL
P4368-11 - 1	LOQ-WATER-05-QT4-2024	925	0.2239	-245.1		PPB	10 Dec 2024 11:14:34	U	SMPL
P5051-07 - 1	14B-(1-4)-COMP	4142	1.1079	-245.1		PPB	10 Dec 2024 11:16:51	U	SMPL
P5068-07 - 1	14B-(1-4)-COMP	3264	0.8667	-245.1		PPB	10 Dec 2024 11:19:08	U	SMPL
P5192-02 - 1	EFF-WASTE WATER	98	-0.0033	-245.1		PPB	10 Dec 2024 11:21:26	U	SMPL
CCV11 - 1	CCV11	17194	4.6946	-245.1		PPB	10 Dec 2024 11:26:17	U	SMPL
CCB11 - 1	CCB11	6	-0.0286	-245.1		PPB	10 Dec 2024 11:31:06	U	SMPL
P5192-02DUP - 1	EFF-WASTE WATERDUP	73	-0.0102	-245.1		PPB	10 Dec 2024 11:33:20	U	SMPL
P5192-02MS - 1	EFF-WASTE WATERMS	13290	3.6218	-245.1		PPB	10 Dec 2024 11:35:35	U	SMPL
P5192-02MSD - 1	EFF-WASTE WATERMSD	14418	3.9318	-245.1		PPB	10 Dec 2024 11:37:51	U	SMPL
P5192-02LX5 - 1		-187	-0.0816	-245.1		PPB	10 Dec 2024 11:40:09	U	SMPL
P5192-02A - 1		13463	3.6693	-245.1		PPB	10 Dec 2024 11:42:27	U	SMPL
CCV12 - 1	CCV12	19587	5.3522	-245.1		PPB	10 Dec 2024 11:44:43	U	SMPL
CCB12 - 1	CCB12	2	-0.0297	-245.1		PPB	10 Dec 2024 11:49:38	U	SMPL