

Surrogate Summary

SDG No.: **P5051**

Client: **New York City DEP of Environmental Protection/BWS**

Analytical Method: **SW624.1**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5051-01	14B-1	1,2-Dichloroethane-d4	30	29.8	99	91	110
		Toluene-d8	30	28.7	96	91	112
		4-Bromofluorobenzene	30	25.0	83	63	112
P5051-02	14B-2	1,2-Dichloroethane-d4	30	30.5	102	91	110
		Toluene-d8	30	29.1	97	91	112
		4-Bromofluorobenzene	30	23.7	79	63	112
P5051-03	14B-3	1,2-Dichloroethane-d4	30	29.6	99	91	110
		Toluene-d8	30	28.6	95	91	112
		4-Bromofluorobenzene	30	23.8	79	63	112
P5051-04	14B-4	1,2-Dichloroethane-d4	30	30.9	103	91	110
		Toluene-d8	30	28.7	96	91	112
		4-Bromofluorobenzene	30	23.8	79	63	112
VN1206WBL01	VN1206WBL01	1,2-Dichloroethane-d4	30	30.7	102	91	110
		Toluene-d8	30	29.0	97	91	112
		4-Bromofluorobenzene	30	24.3	81	63	112
VN1206WBS01	VN1206WBS01	1,2-Dichloroethane-d4	30	30.1	100	91	110
		Toluene-d8	30	29.7	99	91	112
		4-Bromofluorobenzene	30	30.0	100	63	112
VN1206WBSD0	VN1206WBSD01	1,2-Dichloroethane-d4	30	30.8	103	91	110
		Toluene-d8	30	29.6	99	91	112
		4-Bromofluorobenzene	30	29.6	99	63	112