

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133990 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV006				ID: CCV049				ID: CCV050		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	97	1	400000	400000	99	1	390000	97	1
Antimony	1000	940	94	0	5000	5100	101	0	4800	96	1
Arsenic	1000	1100	107	0	5000	5100	102	0	5000	99	1
Barium	520	520	100	0	10000	10000	100	1	9800	98	1
Beryllium	510	530	104	1	500	510	101	1	500	101	1
Cadmium	510	500	98	0	2500	2500	99	0	2500	100	1
Calcium	10000	10000	102	1	400000	400000	100	1	390000	98	1
Chromium	520	550	106	0	15500	15000	99	0	15000	99	1
Cobalt	520	500	97	0	2500	2500	100	0	2500	100	1
Copper	510	510	100	1	15000	15000	99	1	15000	97	1
Iron	10000	10000	103	1	400000	410000	101	1	400000	99	1
Lead	1000	970	97	0	25000	25000	98	0	24000	98	1
Magnesium	6000	6000	100	1	400000	400000	100	1	390000	98	1
Manganese	520	530	102	1	15000	15000	102	1	15000	98	1
Nickel	530	510	96	1	2500	2500	100	0	2500	99	1
Potassium	9900	9700	98	1	150000	150000	101	1	150000	99	1
Selenium	1000	920	92	1	5000	5100	102	0	4900	97	1
Silver	250	240	98	0	1250	1300	102	0	1300	102	1
Sodium	10000	10000	105	3	400000	400000	100	1	380000	96	1
Thallium	1000	1000	103	1	5000	4800	96	0	4900	97	1
Vanadium	500	520	104	1	2500	2500	100	2	2500	100	1
Zinc	1000	960	96	0	15000	16000	106	0	16000	109	2

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133990 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV051				ID: CCV052		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	1	400000	100	0
Antimony					5000	5000	101	1	5000	101	1
Arsenic					5000	5100	101	0	5100	102	0
Barium					10000	10000	100	1	10000	102	0
Beryllium					500	500	100	1	520	103	0
Cadmium					2500	2500	99	1	2500	100	1
Calcium					400000	400000	100	1	410000	101	0
Chromium					15500	15000	99	0	16000	101	1
Cobalt					2500	2500	100	1	2500	100	1
Copper					15000	15000	98	1	15000	101	1
Iron					400000	410000	101	1	410000	103	0
Lead					25000	24000	98	1	25000	99	1
Magnesium					400000	400000	100	1	410000	102	1
Manganese					15000	15000	101	1	15000	103	0
Nickel					2500	2500	99	1	2500	100	1
Potassium					150000	150000	100	1	150000	103	1
Selenium					5000	5100	101	1	5100	101	1
Silver					1250	1300	103	1	1300	103	1
Sodium					400000	390000	99	1	400000	100	1
Thallium					5000	4900	98	0	5000	100	1
Vanadium					2500	2500	99	1	2500	102	1
Zinc					15000	16000	104	1	15000	102	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB133990 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV053				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	390000	98	1			
Antimony					5000	5100	103	1			
Arsenic					5000	5100	103	0			
Barium					10000	10000	102	1			
Beryllium					500	520	103	0			
Cadmium					2500	2500	100	0			
Calcium					400000	410000	102	1			
Chromium					15500	16000	100	1			
Cobalt					2500	2500	100	0			
Copper					15000	15000	101	1			
Iron					400000	400000	101	1			
Lead					25000	25000	99	1			
Magnesium					400000	400000	101	1			
Manganese					15000	16000	104	1			
Nickel					2500	2500	100	0			
Potassium					150000	160000	103	1			
Selenium					5000	5200	103	0			
Silver					1250	1300	103	1			
Sodium					400000	390000	98	1			
Thallium					5000	5000	100	1			
Vanadium					2500	2500	101	1			
Zinc					15000	15000	103	2			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134161 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV051				ID: CCV100				ID: CCV101		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	96	0	400000	410000	102	1	410000	103	1
Antimony	1000	910	91	0	5000	4900	99	0	5000	101	0
Arsenic	1000	1100	108	1	5000	5000	100	0	5000	100	0
Barium	520	520	100	0	10000	10000	103	0	10000	103	0
Beryllium	510	530	104	1	500	520	103	0	520	104	1
Cadmium	510	500	98	0	2500	2500	100	0	2500	102	0
Calcium	10000	10000	102	1	400000	410000	104	0	410000	103	0
Chromium	520	550	106	0	15500	16000	102	0	16000	103	0
Cobalt	520	510	97	0	2500	2500	101	0	2600	103	0
Copper	510	500	97	4	15000	15000	103	0	16000	104	0
Iron	10000	10000	103	1	400000	410000	103	0	420000	104	0
Lead	1000	980	98	0	25000	25000	100	0	25000	101	0
Magnesium	6000	6200	104	2	400000	410000	103	0	420000	104	1
Manganese	520	540	103	1	15000	15000	103	1	16000	103	1
Nickel	530	520	97	0	2500	2500	101	0	2600	102	0
Potassium	9900	9200	93	2	150000	160000	105	0	160000	106	0
Selenium	1000	950	95	1	5000	5000	99	0	5000	100	1
Silver	250	250	99	1	1250	1300	102	0	1300	102	0
Sodium	10000	10000	104	1	400000	420000	104	0	420000	105	1
Thallium	1000	1000	100	1	5000	4900	98	0	5000	99	0
Vanadium	500	510	103	3	2500	2600	104	0	2600	104	1
Zinc	1000	950	95	0	15000	15000	100	0	15000	101	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134161 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV102				ID: CCV103		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	430000	108	7	400000	101	1
Antimony					5000	5000	100	2	4900	98	0
Arsenic					5000	5000	100	0	4900	98	0
Barium					10000	11000	106	7	10000	101	0
Beryllium					500	540	108	6	510	103	1
Cadmium					2500	2500	101	2	2500	99	0
Calcium					400000	430000	107	7	410000	101	0
Chromium					15500	16000	101	1	15000	100	0
Cobalt					2500	2600	102	2	2500	100	0
Copper					15000	16000	108	7	15000	102	0
Iron					400000	430000	109	7	410000	101	1
Lead					25000	25000	101	2	25000	99	0
Magnesium					400000	430000	108	7	410000	102	1
Manganese					15000	16000	108	7	15000	101	1
Nickel					2500	2500	102	2	2500	100	0
Potassium					150000	160000	108	7	150000	103	1
Selenium					5000	5000	99	2	4900	98	0
Silver					1250	1300	101	1	1200	100	0
Sodium					400000	440000	110	6	420000	104	0
Thallium					5000	4900	98	2	4900	97	0
Vanadium					2500	2700	109	7	2500	101	1
Zinc					15000	15000	100	1	15000	98	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134161 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV104				ID: CCV105		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	410000	102	1	400000	101	0
Antimony					5000	5000	100	0	4900	98	0
Arsenic					5000	5100	102	0	5000	100	0
Barium					10000	10000	104	0	10000	103	0
Beryllium					500	530	105	0	520	104	0
Cadmium					2500	2500	102	0	2500	100	0
Calcium					400000	420000	104	1	410000	102	0
Chromium					15500	16000	102	1	16000	100	0
Cobalt					2500	2600	102	0	2500	101	0
Copper					15000	16000	104	1	15000	103	0
Iron					400000	410000	103	0	410000	102	0
Lead					25000	25000	101	0	25000	100	0
Magnesium					400000	410000	103	0	410000	101	0
Manganese					15000	15000	103	0	15000	102	0
Nickel					2500	2600	103	0	2500	101	0
Potassium					150000	160000	104	1	150000	103	1
Selenium					5000	5000	100	0	4900	98	0
Silver					1250	1300	102	0	1300	101	0
Sodium					400000	420000	105	1	410000	102	0
Thallium					5000	5000	99	0	4900	98	0
Vanadium					2500	2600	103	1	2500	102	1
Zinc					15000	15000	100	1	15000	98	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134161 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV106				ID: CCV107		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	100	0	400000	100	0
Antimony					5000	4900	98	0	4900	98	0
Arsenic					5000	5000	100	0	5000	100	0
Barium					10000	10000	102	0	10000	102	0
Beryllium					500	510	102	1	520	103	0
Cadmium					2500	2500	100	0	2500	100	0
Calcium					400000	410000	102	0	410000	102	0
Chromium					15500	15000	100	0	15000	99	0
Cobalt					2500	2500	100	0	2500	100	0
Copper					15000	15000	102	0	15000	102	0
Iron					400000	400000	101	0	410000	102	0
Lead					25000	25000	99	0	25000	99	0
Magnesium					400000	400000	101	0	410000	101	0
Manganese					15000	15000	101	0	15000	101	0
Nickel					2500	2500	101	0	2500	100	0
Potassium					150000	150000	101	1	150000	101	0
Selenium					5000	4900	99	0	4900	98	0
Silver					1250	1300	101	0	1200	100	0
Sodium					400000	410000	102	0	410000	104	1
Thallium					5000	4900	97	0	4800	97	0
Vanadium					2500	2500	101	0	2500	101	1
Zinc					15000	15000	97	0	15000	97	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134161 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV108				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	410000	101	0			
Antimony					5000	4900	98	0			
Arsenic					5000	5000	99	0			
Barium					10000	10000	101	0			
Beryllium					500	520	103	1			
Cadmium					2500	2500	100	0			
Calcium					400000	410000	102	0			
Chromium					15500	16000	100	0			
Cobalt					2500	2500	101	0			
Copper					15000	15000	102	0			
Iron					400000	410000	103	0			
Lead					25000	25000	99	0			
Magnesium					400000	410000	103	1			
Manganese					15000	15000	102	0			
Nickel					2500	2500	101	0			
Potassium					150000	150000	101	0			
Selenium					5000	4900	98	0			
Silver					1250	1200	100	0			
Sodium					400000	420000	106	1			
Thallium					5000	4900	98	0			
Vanadium					2500	2500	102	1			
Zinc					15000	15000	99	0			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHKZ7Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134162 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV052				ID: CCV112				ID: CCV113		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	97	1	400000	420000	104	1	410000	102	0
Antimony	1000	910	91	1	5000	5000	101	0	5000	99	0
Arsenic	1000	1100	107	0	5000	5100	102	0	5000	101	0
Barium	520	510	98	1	10000	10000	105	1	10000	104	0
Beryllium	510	530	103	1	500	520	105	1	520	103	1
Cadmium	510	500	98	0	2500	2600	103	0	2500	101	0
Calcium	10000	10000	102	1	400000	420000	105	1	410000	104	0
Chromium	520	550	106	0	15500	16000	104	0	16000	102	0
Cobalt	520	510	97	0	2500	2600	103	0	2500	102	0
Copper	510	490	95	2	15000	16000	105	1	15000	103	0
Iron	10000	10000	104	1	400000	420000	105	0	410000	103	1
Lead	1000	990	99	0	25000	25000	102	0	25000	100	0
Magnesium	6000	6100	102	1	400000	420000	105	1	410000	102	0
Manganese	520	530	102	1	15000	16000	104	1	15000	102	1
Nickel	530	520	97	0	2500	2600	103	0	2500	102	0
Potassium	9900	9300	94	2	150000	160000	107	1	160000	105	1
Selenium	1000	970	97	1	5000	5100	101	0	5000	100	0
Silver	250	250	99	1	1250	1300	104	0	1300	102	0
Sodium	10000	10000	104	1	400000	430000	107	1	420000	105	1
Thallium	1000	1000	101	1	5000	5000	99	0	4900	98	0
Vanadium	500	510	102	4	2500	2600	104	1	2500	102	0
Zinc	1000	950	95	0	15000	15000	102	1	15000	99	0