

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P5065 SAS No.: P5065 SDG No.: P5065

Instrument ID: BNA_N Calibration Date/Time: 12/03/2024 21:11

Lab File ID: BN035413.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.609		-2.7	50.0
Fluoranthene-d10	1.134	1.023		-9.8	50.0
2-Fluorophenol	1.001	0.907		-9.4	50.0
Phenol-d6	1.204	1.095		-9.1	50.0
Nitrobenzene-d5	0.244	0.240		-1.6	50.0
2-Fluorobiphenyl	1.512	1.547		2.3	50.0
2,4,6-Tribromophenol	0.284	0.256		-9.9	50.0
Terphenyl-d14	0.789	0.795		0.8	50.0
1,4-Dioxane	0.382	0.384		0.5	50.0

All other compounds must meet a minimum RRF of 0.010.