

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28M0

Initial Calibration Verification Source : MP83592

Continuing Calibration Verification Source : MP83594

Run Batch: LB133907 Analytical Method: CVAA

Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV017				ID: CCV020				ID: CCV021		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.2	105		5.0	4.9	98		5.0	99	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28M0

Initial Calibration Verification Source : MP83592

Continuing Calibration Verification Source : MP83594

Run Batch: LB133907 Analytical Method: CVAA

Concentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV022				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	4.7	94				