

**Instrument ID:** P8

**Daily Analysis Runlog For Sequence/QC Batch ID # LB133874**

|              |                 |              |                        |
|--------------|-----------------|--------------|------------------------|
| Review By    | Mohan Bera      | Review On    | 12/11/2024 3:44:01 AM  |
| Supervise By | Sarabjit Jaswal | Supervise On | 12/11/2024 11:09:19 PM |

| STD. NAME     | STD REF.#                                                               |
|---------------|-------------------------------------------------------------------------|
| ICAL Standard | MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016 |
| ICV Standard  | MP83041                                                                 |
| CCV Standard  | MP83042                                                                 |
| ICSA Standard | MP83043,MP83044                                                         |
| CRI Standard  |                                                                         |
| LCS Standard  | MP83465                                                                 |
| Chk Standard  | MP83050,MP83049                                                         |

| Sr# | SampleId   | ClientID | QcType | Date           | Comment                             | Operator | Status   |
|-----|------------|----------|--------|----------------|-------------------------------------|----------|----------|
| 1   | TUNE       | TUNE001  | TUNE   | 12/10/24 10:30 |                                     | Jaswal   | OK       |
| 2   | S00        | S0       | CAL1   | 12/10/24 11:14 |                                     | Jaswal   | OK       |
| 3   | S02        | S02      | CAL3   | 12/10/24 11:21 |                                     | Jaswal   | OK       |
| 4   | S03        | S03      | CAL4   | 12/10/24 11:24 |                                     | Jaswal   | OK       |
| 5   | S04        | S04      | CAL5   | 12/10/24 11:27 |                                     | Jaswal   | OK       |
| 6   | S05        | S05      | CAL6   | 12/10/24 11:30 |                                     | Jaswal   | OK       |
| 7   | S06        | S06      | CAL7   | 12/10/24 11:33 |                                     | Jaswal   | OK       |
| 8   | S07        | S07      | CAL8   | 12/10/24 11:36 |                                     | Jaswal   | OK       |
| 9   | S08        | S08      | CAL9   | 12/10/24 11:38 |                                     | Jaswal   | OK       |
| 10  | ICV001     | ICV001   | ICV    | 12/10/24 13:26 | Inst. paused for calibration review | Jaswal   | OK       |
| 11  | ICB001     | ICB001   | ICB    | 12/10/24 13:30 |                                     | Jaswal   | OK       |
| 12  | ICSA001    | ICSA001  | ICSA   | 12/10/24 13:33 |                                     | Jaswal   | OK       |
| 13  | ICSAB001   | ICSAB001 | ICSAB  | 12/10/24 13:36 |                                     | Jaswal   | OK       |
| 14  | CCV001     | CCV001   | CCV    | 12/10/24 13:39 |                                     | Jaswal   | OK       |
| 15  | CCB001     | CCB001   | CCB    | 12/10/24 14:04 | Inst. paused for review             | Jaswal   | OK       |
| 16  | PB165414BL | PBS414   | MB     | 12/10/24 14:13 |                                     | Jaswal   | OK       |
| 17  | PB165414BS | LCS414   | LCS    | 12/10/24 14:25 |                                     | Jaswal   | OK       |
| 18  | P5067-01   | ME28M0   | SAM    | 12/10/24 14:28 | Cr,Mn,V High                        | Jaswal   | Dilution |

Instrument ID: P8

**Daily Analysis Runlog For Sequence/QC Batch ID # LB133874**

|                  |                                                                         |              |                        |
|------------------|-------------------------------------------------------------------------|--------------|------------------------|
| Review By        | Mohan Bera                                                              | Review On    | 12/11/2024 3:44:01 AM  |
| Supervise By     | Sarabjit Jaswal                                                         | Supervise On | 12/11/2024 11:09:19 PM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                                        |              |                        |
| ICAL Standard    | MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016 |              |                        |
| ICV Standard     | MP83041                                                                 |              |                        |
| CCV Standard     | MP83042                                                                 |              |                        |
| ICSA Standard    | MP83043,MP83044                                                         |              |                        |
| CRI Standard     |                                                                         |              |                        |
| LCS Standard     | MP83465                                                                 |              |                        |
| Chk Standard     | MP83050,MP83049                                                         |              |                        |

|    |          |        |     |                |            |        |          |
|----|----------|--------|-----|----------------|------------|--------|----------|
| 19 | P5067-02 | ME28M1 | SAM | 12/10/24 14:31 | Cr,Mn High | Jaswal | Dilution |
| 20 | P5067-03 | ME28M2 | SAM | 12/10/24 14:34 | Mn High    | Jaswal | Dilution |
| 21 | P5067-04 | ME28M3 | SAM | 12/10/24 14:38 |            | Jaswal | OK       |
| 22 | P5067-05 | ME28M4 | SAM | 12/10/24 14:41 |            | Jaswal | OK       |
| 23 | CCV002   | CCV002 | CCV | 12/10/24 14:44 |            | Jaswal | OK       |
| 24 | CCB002   | CCB002 | CCB | 12/10/24 14:47 |            | Jaswal | OK       |
| 25 | P5067-06 | ME28M5 | SAM | 12/10/24 14:50 |            | Jaswal | OK       |
| 26 | P5067-07 | ME28M6 | SAM | 12/10/24 14:53 |            | Jaswal | OK       |
| 27 | P5067-08 | ME28M7 | SAM | 12/10/24 14:56 |            | Jaswal | OK       |
| 28 | P5067-09 | ME28M8 | SAM | 12/10/24 14:59 |            | Jaswal | OK       |
| 29 | P5067-10 | ME28M9 | SAM | 12/10/24 15:03 |            | Jaswal | OK       |
| 30 | P5067-11 | ME28N0 | SAM | 12/10/24 15:06 |            | Jaswal | OK       |
| 31 | P5067-12 | ME28N1 | SAM | 12/10/24 15:09 |            | Jaswal | OK       |
| 32 | P5067-13 | ME28N2 | SAM | 12/10/24 15:12 |            | Jaswal | OK       |
| 33 | P5067-14 | ME28N3 | SAM | 12/10/24 15:15 |            | Jaswal | OK       |
| 34 | P5067-15 | ME28N4 | SAM | 12/10/24 15:19 |            | Jaswal | OK       |
| 35 | P5067-16 | ME28N5 | SAM | 12/10/24 15:22 |            | Jaswal | OK       |
| 36 | P5067-17 | ME28N6 | SAM | 12/10/24 15:25 |            | Jaswal | OK       |
| 37 | P5067-18 | ME28N7 | SAM | 12/10/24 15:28 |            | Jaswal | OK       |
| 38 | P5067-19 | ME28N8 | SAM | 12/10/24 15:31 |            | Jaswal | OK       |

Instrument ID: P8

**Daily Analysis Runlog For Sequence/QC Batch ID # LB133874**

|                  |                                                                         |              |                        |
|------------------|-------------------------------------------------------------------------|--------------|------------------------|
| Review By        | Mohan Bera                                                              | Review On    | 12/11/2024 3:44:01 AM  |
| Supervise By     | Sarabjit Jaswal                                                         | Supervise On | 12/11/2024 11:09:19 PM |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                                        |              |                        |
| ICAL Standard    | MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016 |              |                        |
| ICV Standard     | MP83041                                                                 |              |                        |
| CCV Standard     | MP83042                                                                 |              |                        |
| ICSA Standard    | MP83043,MP83044                                                         |              |                        |
| CRI Standard     |                                                                         |              |                        |
| LCS Standard     | MP83465                                                                 |              |                        |
| Chk Standard     | MP83050,MP83049                                                         |              |                        |

|    |             |         |     |                |                              |        |          |
|----|-------------|---------|-----|----------------|------------------------------|--------|----------|
| 39 | P5067-20    | ME28N8D | DUP | 12/10/24 15:34 |                              | Jaswal | OK       |
| 40 | P5067-19L   | ME28N8L | SD  | 12/10/24 15:38 |                              | Jaswal | OK       |
| 41 | P5067-21    | ME28N8S | MS  | 12/10/24 15:41 |                              | Jaswal | OK       |
| 42 | P5067-01DL  | ME28M0  | SAM | 12/10/24 16:09 | 2x for Cr,Mn,V Still Mn high | Jaswal | Dilution |
| 43 | P5067-01DL2 | ME28M0  | SAM | 12/10/24 16:12 | 5x for Mn                    | Jaswal | Confirms |
| 44 | P5067-02DL  | ME28M1  | SAM | 12/10/24 16:15 | 5x for Cr,Mn                 | Jaswal | Confirms |
| 45 | P5067-03DL  | ME28M2  | SAM | 12/10/24 16:18 | 5x for Mn                    | Jaswal | Confirms |
| 46 | CCV003      | CCV003  | CCV | 12/10/24 16:26 |                              | Jaswal | OK       |
| 47 | CCB003      | CCB003  | CCB | 12/10/24 16:30 |                              | Jaswal | OK       |