

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28M0
 Instrument ID: P8 Start Date: 12/10/2024
 Analytical Method: ICP-MS Run Batch: LB133874
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	2.1	5	50	49	-2
Arsenic	0	0.0	0	1	1.0	0	50	53	6
Barium	0	0.0	0	10	10	0	250	250	0
Beryllium	0	0.0	0	1	1.1	10	50	50	0
Cadmium	0	0.0	0	1	1.1	10	50	50	0
Chromium	0	0.0	0	2	2.1	5	50	51	2
Cobalt	0	0.0	0	1	1.1	10	50	53	6
Copper	0	0.0	0	2	2.3	15	500	550	10
Lead	0	0.0	0	1	0.96	-4	250	240	-4
Manganese	0	0.0	0	1	0.98	-2	500	510	2
Nickel	0	0.0	0	1	1.0	0	50	53	6
Selenium	0	0.0	0	5	4.5	-10	50	51	2
Silver	0	0.0	0	1	1.1	10	50	53	6
Thallium	0	0.0	0	1	0.97	-3	50	49	-2
Vanadium	0	0.0	0	5	5.1	2	50	51	2
Zinc	0	0.0	0	5	5.6	12	500	520	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME28M0
 Instrument ID: P8 Start Date: 12/10/2024
 Analytical Method: ICP-MS Run Batch: LB133874
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	130	4	250	250	0	500	510	2
Arsenic	125	140	12	250	260	4	500	500	0
Barium	625	640	2	1250	1200	-4	2500	2500	0
Beryllium	125	130	4	250	260	4	500	510	2
Cadmium	125	130	4	250	250	0	500	510	2
Chromium	125	130	4	250	260	4	500	500	0
Cobalt	125	130	4	250	260	4	500	510	2
Copper	1250	1400	12	2500	2600	4	5000	5100	2
Lead	625	620	-1	1250	1200	-4	2500	2500	0
Manganese	1250	1300	4	2500	2500	0	5000	5000	0
Nickel	125	130	4	250	250	0	500	480	-4
Selenium	125	130	4	250	250	0	500	510	2
Silver	125	140	12	250	260	4	500	520	4
Thallium	125	120	-4	250	240	-4	500	490	-2
Vanadium	125	130	4	250	250	0	500	500	0
Zinc	1250	1300	4	2500	2600	4	5000	5000	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME28M0
Instrument ID: P8 Start Date: 12/10/2024
Analytical Method: ICP-MS Run Batch: LB133874
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	1000	0						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	1000	0						
Cadmium	1000	1000	0						
Chromium	1000	1000	0						
Cobalt	1000	990	-1						
Copper	10000	9900	-1						
Lead	5000	5000	0						
Manganese	10000	10000	0						
Nickel	1000	1000	0						
Selenium	1000	1000	0						
Silver	1000	990	-1						
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	10000	0						