

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28M0Initial Calibration Verification Source : MP83041Continuing Calibration Verification Source : MP83042Run Batch: LB133874 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	220	108	1	500	500	100	1	500	100	0
Arsenic	200	220	109	1	500	540	108	6	500	100	1
Barium	100	100	104	1	2500	2500	101	2	2500	100	2
Beryllium	100	110	107	2	500	520	104	3	520	104	2
Cadmium	100	110	108	1	500	490	99	2	490	98	2
Chromium	100	100	102	0	500	530	106	6	510	101	1
Cobalt	100	110	105	1	500	520	104	5	490	99	1
Copper	100	99	99	0	5000	5000	100	5	4800	95	0
Lead	200	200	102	2	2500	2500	101	1	2500	102	1
Manganese	100	100	100	1	5000	5300	106	6	5000	100	1
Nickel	110	110	96	1	500	480	96	6	450	90	1
Selenium	200	210	104	2	500	490	98	2	490	97	2
Silver	50.0	55	109	1	500	480	97	3	480	97	1
Thallium	210	200	97	2	500	510	103	1	520	103	2
Vanadium	100	99	99	1	500	540	109	6	510	102	1
Zinc	200	200	99	1	5000	5000	100	6	4700	94	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28M0Initial Calibration Verification Source : MP83041Continuing Calibration Verification Source : MP83042Run Batch: LB133874 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	490	98	2			
Arsenic					500	490	99	1			
Barium					2500	2500	99	2			
Beryllium					500	510	101	2			
Cadmium					500	480	97	1			
Chromium					500	510	101	1			
Cobalt					500	490	99	1			
Copper					5000	4700	95	1			
Lead					2500	2500	100	2			
Manganese					5000	4900	98	1			
Nickel					500	450	90	1			
Selenium					500	470	95	2			
Silver					500	480	95	1			
Thallium					500	510	101	2			
Vanadium					500	510	102	3			
Zinc					5000	4800	95	2			