

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB133874

Review By	Mohan Bera	Review On	12/11/2024 3:44:01 AM
Supervise By	Sarabjit Jaswal	Supervise On	12/11/2024 11:09:19 PM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83465
Chk Standard	MP83050,MP83049

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	12/10/24 10:30		Jaswal	OK
2	S00	S0	CAL1	12/10/24 11:14		Jaswal	OK
3	S02	S02	CAL3	12/10/24 11:21		Jaswal	OK
4	S03	S03	CAL4	12/10/24 11:24		Jaswal	OK
5	S04	S04	CAL5	12/10/24 11:27		Jaswal	OK
6	S05	S05	CAL6	12/10/24 11:30		Jaswal	OK
7	S06	S06	CAL7	12/10/24 11:33		Jaswal	OK
8	S07	S07	CAL8	12/10/24 11:36		Jaswal	OK
9	S08	S08	CAL9	12/10/24 11:38		Jaswal	OK
10	ICV001	ICV001	ICV	12/10/24 13:26	Inst. paused for calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	12/10/24 13:30		Jaswal	OK
12	ICSA001	ICSA001	ICSA	12/10/24 13:33		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	12/10/24 13:36		Jaswal	OK
14	CCV001	CCV001	CCV	12/10/24 13:39		Jaswal	OK
15	CCB001	CCB001	CCB	12/10/24 14:04	Inst. paused for review	Jaswal	OK
16	PB165414BL	PBS414	MB	12/10/24 14:13		Jaswal	OK
17	PB165414BS	LCS414	LCS	12/10/24 14:25		Jaswal	OK
18	P5067-01	ME28M0	SAM	12/10/24 14:28	Cr,Mn,V High	Jaswal	Dilution

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB133874

Review By	Mohan Bera	Review On	12/11/2024 3:44:01 AM
Supervise By	Sarabjit Jaswal	Supervise On	12/11/2024 11:09:19 PM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83465
Chk Standard	MP83050,MP83049

19	P5067-02	ME28M1	SAM	12/10/24 14:31	Cr,Mn High	Jaswal	Dilution
20	P5067-03	ME28M2	SAM	12/10/24 14:34	Mn High	Jaswal	Dilution
21	P5067-04	ME28M3	SAM	12/10/24 14:38		Jaswal	OK
22	P5067-05	ME28M4	SAM	12/10/24 14:41		Jaswal	OK
23	CCV002	CCV002	CCV	12/10/24 14:44		Jaswal	OK
24	CCB002	CCB002	CCB	12/10/24 14:47		Jaswal	OK
25	P5067-06	ME28M5	SAM	12/10/24 14:50		Jaswal	OK
26	P5067-07	ME28M6	SAM	12/10/24 14:53		Jaswal	OK
27	P5067-08	ME28M7	SAM	12/10/24 14:56		Jaswal	OK
28	P5067-09	ME28M8	SAM	12/10/24 14:59		Jaswal	OK
29	P5067-10	ME28M9	SAM	12/10/24 15:03		Jaswal	OK
30	P5067-11	ME28N0	SAM	12/10/24 15:06		Jaswal	OK
31	P5067-12	ME28N1	SAM	12/10/24 15:09		Jaswal	OK
32	P5067-13	ME28N2	SAM	12/10/24 15:12		Jaswal	OK
33	P5067-14	ME28N3	SAM	12/10/24 15:15		Jaswal	OK
34	P5067-15	ME28N4	SAM	12/10/24 15:19		Jaswal	OK
35	P5067-16	ME28N5	SAM	12/10/24 15:22		Jaswal	OK
36	P5067-17	ME28N6	SAM	12/10/24 15:25		Jaswal	OK
37	P5067-18	ME28N7	SAM	12/10/24 15:28		Jaswal	OK
38	P5067-19	ME28N8	SAM	12/10/24 15:31		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB133874

Review By	Mohan Bera	Review On	12/11/2024 3:44:01 AM
Supervise By	Sarabjit Jaswal	Supervise On	12/11/2024 11:09:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83465		
Chk Standard	MP83050,MP83049		

39	P5067-20	ME28N8D	DUP	12/10/24 15:34		Jaswal	OK
40	P5067-19L	ME28N8L	SD	12/10/24 15:38		Jaswal	OK
41	P5067-21	ME28N8S	MS	12/10/24 15:41		Jaswal	OK
42	P5067-01DL	ME28M0	SAM	12/10/24 16:09	2x for Cr,Mn,V Still Mn high	Jaswal	Dilution
43	P5067-01DL2	ME28M0	SAM	12/10/24 16:12	5x for Mn	Jaswal	Confirms
44	P5067-02DL	ME28M1	SAM	12/10/24 16:15	5x for Cr,Mn	Jaswal	Confirms
45	P5067-03DL	ME28M2	SAM	12/10/24 16:18	5x for Mn	Jaswal	Confirms
46	CCV003	CCV003	CCV	12/10/24 16:26		Jaswal	OK
47	CCB003	CCB003	CCB	12/10/24 16:30		Jaswal	OK