

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. P5068
QUOTE NO. _____
COC Number 2042253

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: New York City Dep of Environmental

ADDRESS: 3701 Jerome Ave

CITY Bronx STATE: NY ZIP: _____

ATTENTION: Nicholas Prokopovics

PHONE: _____

FAX: _____

CLIENT PROJECT INFORMATION

PROJECT NAME: Industrial w/w Discharge Permit 2024

PROJECT NO.: _____ LOCATION: _____

PROJECT MANAGER: _____

e-mail: _____

PHONE: _____

FAX: _____

CLIENT BILLING INFORMATION

BILL TO: _____ PO#: _____

ADDRESS: _____

CITY _____ STATE: _____ ZIP: _____

ATTENTION: _____ PHONE: _____

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*

HARDCOPY (DATA PACKAGE): _____ DAYS*

EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1 VOCs 2 Metals 3 Cyanide 4 OB-G 5 TSS, H&T, Chlorophyll 6 7 8 9

PRESERVATIVES

COMMENTS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		A	B	D	C	E					
1.	14B#1	AQ		X	12-03-24	0740	6	X	X	X	X	X					6.62 / 11.01°C
2.	14B#2			X		0840	6	X	X	X	X	X					6.66 / 12.15°C
3.	14B#3			X		0940	6	X	X	X	X	X					6.57 / 12.30°C
4.	14B#4			X		1040	6	X	X	X	X	X					6.51 / 11.83°C
5.	14B#4 MS/MSD			X		1042	2				X						
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>1130</u> 12-03-24	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP <u>2.6</u> °C
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME: _____	RECEIVED BY: 2. <u>[Signature]</u>	Comments: _____
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>1350</u> 12-03-24	RECEIVED BY: 3. <u>[Signature]</u>	Page <u>1</u> of <u>1</u>

CLIENT: ☐ Hand Delivered ☐ Other _____
CHEMTECH: ☐ Picked Up ☒ Field Sampling

Shipment Complete
☒ YES ☐ NO

284 Sheffield Street, Mountainside, NJ 07092 Tel. 908-789-8900 Fax 908-789-8922
FIELD SAMPLING LOG

Project Name: Industrial w/w Discharge permit

Project Location: Brooklyn NY

Cooler Custody Seal: *N/A*

Temperature Correction Factor (°C): *N/A*

Departure Time: 11 30

FIELD SAMPLING INFORMATION

[illegible]

Meter: YSI MPS, Model # 556, Serial # 085A0063

Sampler Signature/Date:

Supervisor Review/Date:

Chemtech Project number: P5068

Date: 12-03-24

Client Name: New York Dep of Environmental

Client Project Name : Industrial w/w Discharge permit 2024

Instructions: Composite metals, Cyanide, Hex chrom samples 4:1

Sample Custodian: JM

QA Control # A3041240

Alliance Technical Group, LLC-Newark

284 Sheffield Street, Mountainside, NJ 07092 Tel. 908-789-8900 Fax 908-789-8922

FIELD SAMPLING LOG

Client Name: New York City Dep of Environmental

Client Address: 3701 Jerome Ave, Bronx, NY

Client Rep on Site: Nicholas Prokopowicz

Sampling Date: 12-03-24

Arrival Time: 0658

Departure Time: 1130

Project Name: Industrial w/w discharge permit 2024

Project Location: Bronx NY

Cooler Custody Seal: N/A

Temperature Correction Factor (°C): N/A

FIELD EQUIPMENT CALIBRATION

pH Calibration (SM4500-H B/9040C)				
Calibration				ICV (± 0.1 pH unit)
	7.00 Buffer W 3071	4.00 Buffer W 3107	10.00 Buffer W 3094	7.00 Buffer W 3093
Time	0710	0716	0720	0725
Temp °C	13.06	12.89	12.63	12.28
pH	6.99	3.99	9.98	7.01

FIELD EQUIPMENT CALIBRATION

Specific Conductance (mS/cm) (99% -101%)/(mmho/cm) (SM2510 B/120.1/9050A)		
Calibration (± 1%) (99% -101%)		ICV (± 1%) (99% -101%)
	WP	WP
Time		
Temp °C		
Reading (mS/cm)		

Sampler Signature/Date: 2M 12-03-24

Supervisor Review/Date: [Signature] 12-3-24