

Surrogate Summary

SDG No.: P5068

Client: New York City DEP of Environmental Protection/BWS

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5068-01	14B-1	1,2-Dichloroethane-d4	30	29.8	99	91	110
		Toluene-d8	30	28.9	96	91	112
		4-Bromofluorobenzene	30	24.0	80	63	112
P5068-02	14B-2	1,2-Dichloroethane-d4	30	29.7	99	91	110
		Toluene-d8	30	28.7	96	91	112
		4-Bromofluorobenzene	30	24.1	80	63	112
P5068-03	14B-3	1,2-Dichloroethane-d4	30	30.2	101	91	110
		Toluene-d8	30	29.0	97	91	112
		4-Bromofluorobenzene	30	23.5	78	63	112
P5068-04	14B-4	1,2-Dichloroethane-d4	30	29.9	100	91	110
		Toluene-d8	30	28.3	94	91	112
		4-Bromofluorobenzene	30	23.4	78	63	112
VN1206WBL01	VN1206WBL01	1,2-Dichloroethane-d4	30	30.7	102	91	110
		Toluene-d8	30	29.0	97	91	112
		4-Bromofluorobenzene	30	24.3	81	63	112
VN1206WBS01	VN1206WBS01	1,2-Dichloroethane-d4	30	30.1	100	91	110
		Toluene-d8	30	29.7	99	91	112
		4-Bromofluorobenzene	30	30.0	100	63	112
VN1206WBSD0	VN1206WBSD01	1,2-Dichloroethane-d4	30	30.8	103	91	110
		Toluene-d8	30	29.6	99	91	112
		4-Bromofluorobenzene	30	29.6	99	63	112