


INSTRUMENT ID: CV2



LB133780 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Dec 2024 11:06:52

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.000	7470A	PPB	06 Dec 2024 11:18:38	S	Std
0.2 - 1	50.2	861	-	0.200	7470A	PPB	06 Dec 2024 11:20:55	S	Std
2.5 - 1	52.5	8203	-	2.500	7470A	PPB	06 Dec 2024 11:23:12	S	Std
5.0 - 1	55	16853	-	5.000	7470A	PPB	06 Dec 2024 11:25:29	S	Std
7.5 - 1	57.5	25454	-	7.500	7470A	PPB	06 Dec 2024 11:30:25	S	Std
10.0 - 1	58	33222	-	10.000	7470A	PPB	06 Dec 2024 11:35:22	S	Std
ICV11 - 1	ICV11	12914	3.8420	-	7470A	PPB	06 Dec 2024 11:38:18	U	SMPL
ICB11 - 1	ICB11	-204	-0.0895	-	7470A	PPB	06 Dec 2024 11:40:33	U	SMPL
CCV95 - 1	CCV95	16787	5.0027	-	7470A	PPB	06 Dec 2024 11:42:49	U	SMPL
CCB95 - 1	CCB95	-177	-0.0815	-	7470A	PPB	06 Dec 2024 11:45:04	U	SMPL
CRA - 1	CRA	731	0.1907	-	7470A	PPB	06 Dec 2024 11:47:21	U	SMPL
HighStd - 1	HighStd	33332	9.9613	-	7470A	PPB	06 Dec 2024 11:49:36	U	SMPL
ChkStd - 1	ChkStd	22782	6.7994	-	7470A	PPB	06 Dec 2024 11:51:52	U	SMPL
PB165437BL - 1	PBW	-36	-0.0392	-	7470A	PPB	06 Dec 2024 11:56:55	U	SMPL
PB165437BS - 1	LCSW	16026	4.7746	-	7470A	PPB	06 Dec 2024 11:59:10	U	SMPL
P4368-07 - 1	LOD-MDL-WATER-01-QT4-2024	302	0.0621	-	7470A	PPB	06 Dec 2024 12:12:11	U	SMPL
P4368-08 - 1	LOQ-WATER-02-QT4-2024	878	0.2347	-	7470A	PPB	06 Dec 2024 12:14:27	U	SMPL
P4813-07 - 1	419	1177	0.3243	-	7470A	PPB	06 Dec 2024 12:16:45	U	SMPL
P5074-02 - 1	COMP	1095	0.2998	-	7470A	PPB	06 Dec 2024 12:23:56	U	SMPL
P5074-02DUP - 1	COMPDUP	1040	0.2833	-	7470A	PPB	06 Dec 2024 12:26:11	U	SMPL
CCV96 - 1	CCV96	18126	5.4040	-	7470A	PPB	06 Dec 2024 12:36:13	U	SMPL
CCB96 - 1	CCB96	-60	-0.0464	-	7470A	PPB	06 Dec 2024 12:41:03	U	SMPL
P5074-02MS - 1	COMPMS	10140	3.0108	-	7470A	PPB	06 Dec 2024 12:43:18	U	SMPL
P5074-02MSD - 1	COMPMSD	10105	3.0001	-	7470A	PPB	06 Dec 2024 12:45:33	U	SMPL
P5093-01 - 1	LL-001	-368	-0.1387	-	7470A	PPB	06 Dec 2024 12:47:51	U	SMPL
P5093-02 - 1	LL-001-FB-12-4-24	-68	-0.0488	-	7470A	PPB	06 Dec 2024 12:50:10	U	SMPL
P5145-01 - 1	286085	843	0.2242	-	7470A	PPB	06 Dec 2024 12:52:26	U	SMPL
P5074-02LX5 - 1		-164	-0.0776	-	7470A	PPB	06 Dec 2024 12:54:42	U	SMPL
P5074-02A - 1		11492	3.4158	-	7470A	PPB	06 Dec 2024 12:57:01	U	SMPL
CCV97 - 1	CCV97	15238	4.5385	-	7470A	PPB	06 Dec 2024 12:59:17	U	SMPL
CCB97 - 1	CCB97	-127	-0.0665	-	7470A	PPB	06 Dec 2024 13:01:36	U	SMPL