



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms **SDG No.:** P5074
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5074 **SAS No.:** P5074
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	95	40 - 160	CV	12/06/2024	11:47	LB133780
CRI01	Cadmium	5.80	6.0	97	40 - 160	P	12/10/2024	11:24	LB133866
	Copper	22.5	20.0	113	40 - 160	P	12/10/2024	11:24	LB133866
	Lead	11.2	12.0	94	40 - 160	P	12/10/2024	11:24	LB133866
	Nickel	39.3	40.0	98	40 - 160	P	12/10/2024	11:24	LB133866
	Zinc	43.8	40.0	110	40 - 160	P	12/10/2024	11:24	LB133866