



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Aramark Uniforms				SDG No.:	P5074				
Contract:	ARAM01	Lab Code:	CHEM		Case No.:	P5074	SAS No.:	P5074		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB11	Mercury	0.20	+/-0.20	U	0.20	CV	12/06/2024	11:40	LB133780	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: P5074

Contract: ARAM01

Lab Code: CHEM

Case No.: P5074

SAS No.: P5074

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB95	Mercury	0.20	+/-0.20	U	0.20	CV	12/06/2024	11:45	LB133780
CCB96	Mercury	0.20	+/-0.20	U	0.20	CV	12/06/2024	12:41	LB133780
CCB97	Mercury	0.20	+/-0.20	U	0.20	CV	12/06/2024	13:01	LB133780

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: P5074

Contract: ARAM01

Lab Code: CHEM

Case No.: P5074

SAS No.: P5074

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	12/10/2024	11:20	LB133866
	Copper	20.0	+/-20.0	U	20.0	P	12/10/2024	11:20	LB133866
	Lead	12.0	+/-12.0	U	12.0	P	12/10/2024	11:20	LB133866
	Nickel	40.0	+/-40.0	U	40.0	P	12/10/2024	11:20	LB133866
	Zinc	40.0	+/-40.0	U	40.0	P	12/10/2024	11:20	LB133866

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Aramark Uniforms</u>		SDG No.: <u>P5074</u>							
Contract: <u>ARAM01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5074</u>	SAS No.: <u>P5074</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	12/10/2024	11:48	LB133866
	Copper	20.0	+/-20.0	U	20.0	P	12/10/2024	11:48	LB133866
	Lead	12.0	+/-12.0	U	12.0	P	12/10/2024	11:48	LB133866
	Nickel	40.0	+/-40.0	U	40.0	P	12/10/2024	11:48	LB133866
	Zinc	40.0	+/-40.0	U	40.0	P	12/10/2024	11:48	LB133866
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	12/10/2024	12:39	LB133866
	Copper	20.0	+/-20.0	U	20.0	P	12/10/2024	12:39	LB133866
	Lead	12.0	+/-12.0	U	12.0	P	12/10/2024	12:39	LB133866
	Nickel	40.0	+/-40.0	U	40.0	P	12/10/2024	12:39	LB133866
	Zinc	40.0	+/-40.0	U	40.0	P	12/10/2024	12:39	LB133866
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	12/10/2024	13:30	LB133866
	Copper	20.0	+/-20.0	U	20.0	P	12/10/2024	13:30	LB133866
	Lead	12.0	+/-12.0	U	12.0	P	12/10/2024	13:30	LB133866
	Nickel	40.0	+/-40.0	U	40.0	P	12/10/2024	13:30	LB133866
	Zinc	40.0	+/-40.0	U	40.0	P	12/10/2024	13:30	LB133866
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	12/10/2024	14:24	LB133866
	Copper	20.0	+/-20.0	U	20.0	P	12/10/2024	14:24	LB133866
	Lead	12.0	+/-12.0	U	12.0	P	12/10/2024	14:24	LB133866
	Nickel	40.0	+/-40.0	U	40.0	P	12/10/2024	14:24	LB133866
	Zinc	40.0	+/-40.0	U	40.0	P	12/10/2024	14:24	LB133866