

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Aramark Uniforms **level:** low **sdg no.:** P5074
contract: ARAM01 **lab code:** CHEM **case no.:** P5074 **sas no.:** P5074
matrix: Water **sample id:** P5074-02 **client id:** COMPMS
Percent Solids for Sample: NA **Spiked ID:** P5074-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	82 - 119	3.01		0.30		4.0	68	N	CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Aramark Uniforms **level:** low **sdg no.:** P5074
contract: ARAM01 **lab code:** CHEM **case no.:** P5074 **sas no.:** P5074
matrix: Water **sample id:** P5074-02 **client id:** COMPMSD
Percent Solids for Sample: NA **Spiked ID:** P5074-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	82 - 119	3.00		0.30		4.0	68	N	CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>P5074</u>
contract: <u>ARAM01</u>	lab code: <u>CHEM</u>	case no.: <u>P5074</u> sas no.: <u>P5074</u>
matrix: <u>Water</u>	sample id: <u>P5093-02</u>	client id: <u>LL-001-FB-12-4-24MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P5093-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	78.8		3.00	U	100	79		P
Copper	ug/L	75 - 125	134		10.0	U	150	89		P
Lead	ug/L	75 - 125	392		6.00	U	500	78		P
Nickel	ug/L	75 - 125	207		20.0	U	250	83		P
Zinc	ug/L	75 - 125	91.4		20.0	U	100	91		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Aramark Uniforms **level:** low **sdg no.:** P5074
contract: ARAM01 **lab code:** CHEM **case no.:** P5074 **sas no.:** P5074
matrix: Water **sample id:** P5093-02 **client id:** LL-001-FB-12-4-24MSD
Percent Solids for Sample: NA **Spiked ID:** P5093-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	78.4		3.00	U	100	78		P
Copper	ug/L	75 - 125	133		10.0	U	150	89		P
Lead	ug/L	75 - 125	391		6.00	U	500	78		P
Nickel	ug/L	75 - 125	205		20.0	U	250	82		P
Zinc	ug/L	75 - 125	90.3		20.0	U	100	90		P