

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Aramark Uniforms</u>	Level:	<u>LOW</u>	SDG No.:	<u>P5074</u>
Contract:	<u>ARAM01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5074</u>
Matrix:	<u>Water</u>	Sample ID:	<u>P5074-02</u>	Client ID:	<u>COMPDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	P5074-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.30		0.28		6		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Aramark Uniforms **Level:** LOW **SDG No.:** P5074
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5074 **SAS No.:** P5074
Matrix: Water **Sample ID:** P5074-02MS **Client ID:** COMPMSD
Percent Solids for Sample: NA **Duplicate ID** P5074-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.01		3.00		0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Aramark Uniforms **Level:** LOW **SDG No.:** P5074
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5074 **SAS No.:** P5074
Matrix: Water **Sample ID:** P5093-02 **Client ID:** LL-001-FB-12-4-24DUP
Percent Solids for Sample: NA **Duplicate ID** P5093-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Lead	ug/L	20	6.00	U	6.00	U			P
Nickel	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	20.0	U	20.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Aramark Uniforms **Level:** LOW **SDG No.:** P5074
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P5074 **SAS No.:** P5074
Matrix: Water **Sample ID:** P5093-02MS **Client ID:** LL-001-FB-12-4-24MSD
Percent Solids for Sample: NA **Duplicate ID** P5093-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	78.8		78.4		1		P
Copper	ug/L	20	134		133		1		P
Lead	ug/L	20	392		391		0		P
Nickel	ug/L	20	207		205		1		P
Zinc	ug/L	20	91.4		90.3		1		P