

## **DATA PACKAGE GENERAL CHEMISTRY**

**PROJECT NAME : FT MEADE TIPTON AIRFIELD PARCEL RI - PO 0111169**

### **WESTON SOLUTIONS**

**1400 Weston Way**

**PO Box 2653**

**West Chester, PA - 19380**

**Phone No: 610-701-7400**

**ORDER ID : P5076**

**ATTENTION : Nathan Fretz**



**Laboratory Certification ID # 20012**



1) GENERAL CHEMISTRY DATA	2
2) Signature Page	3
3) Case Narrative	4
4) Qualifier Page	5
5) Conformance/Non Conformance	6
6) QA Checklist	7
7) Chronicle	8
8) Sample Data	9
8.1) TAPIAL2-SB02I-7.5-120224-00-T1	10
9) QC Data Summary For Genchem	11
9.1) Initial and Continuing Calibration Verification	12
9.2) Initial and Continuing Calibration Blank Summary	14
9.3) Preparation Blank Summary	15
9.4) Matrix Spike Summary	16
9.5) Duplicate Sample Summary	18
9.6) Laboratory Control Sample Summary	20
10) GENCHEM RAW DATA	21
10.1) GENCHEM RAW DATA - ANALYTICAL	22
10.1.1) LB133750	22
10.1.2) LB133779	24
11) Analytical Runlogs	42
12) Standard Prep Logs	45
13) Percent Solid	69
14) Shipping Document	73
14.1) Chain Of Custody	74
14.2) Lab Certificate	75

## Cover Page

**Order ID :** P5076

**Project ID :** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**Client :** Weston Solutions

**Lab Sample Number**

P5076-01

**Client Sample Number**

TAPIAL2-SB02I-7.5-120224-00-T1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

**APPROVED**

*By Sohil Jodhani, QA/QC Director at 10:48 am, Dec 23, 2024*

Date: 12/10/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Weston Solutions**

**Project Name:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**Project #** N/A

**Chemtech Project #** P5076

**Test Name:** pH,TOC

### **A. Number of Samples and Date of Receipt:**

1 Solid sample was received on 12/04/2024.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Anions Group1, Mercury, Metals ICP-TAL, METALS-TAL, pH and TOC. This data package contains results for pH,TOC.

### **C. Analytical Techniques:**

The analysis of pH was based on method 9045D and The analysis of TOC was based on method 9060A.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for TAPIAL2-SB02I-7.5-120224-00-T1 of pH as sample receive out of holding time.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature

**APPROVED**

*By Sohil Jodhani, QA/QC Director at 10:49 am, Dec 23, 2024*

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

<b>J</b>	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
<b>U</b>	Indicates the analyte was analyzed for, but not detected.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>E</b>	Indicates the reported value is estimated because of the presence of interference
<b>M</b>	Indicates Duplicate injection precision not met.
<b>N</b>	Indicates the spiked sample recovery is not within control limits.
<b>S</b>	Indicates the reported value was determined by the Method of Standard Addition (MSA).
<b>*</b>	Indicates that the duplicate analysis is not within control limits.
<b>+</b>	Indicates the correlation coefficient for the MSA is less than 0.995.
<b>D</b>	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
<b>M</b>	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
<b>OR</b>	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements
<b>H</b>	Sample Analysis Out Of Hold Time

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: P5076

MATRIX: Solid

METHOD: 9045D,9060A

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.  The Blank Spike met requirements for all samples.			✓
3. Sample Duplicate Analysis Met QC Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓
4. Digestion Holding Time Met  If not met, list number of days exceeded for each sample:  The Holding Times were met for all samples except for TAPIAL2-SB02I-7.5-120224-00-T1 of pH as sample receive out of holding time.		✓	

ADDITIONAL COMMENTS:

QA REVIEW

**APPROVED**

By Sohil Jodhani, QA/QC Director at 10:49 am, Dec 23, 2024

APPENDIX A

**QA REVIEW GENERAL DOCUMENTATION**

Project #: P5076

Completed

For thorough review, the report must have the following:

**GENERAL:**

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

**COVER PAGE:**

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

**CHAIN OF CUSTODY:**

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

**ANALYTICAL:**

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 12/10/2024

LAB CHRONICLE

OrderID:	P5076	OrderDate:	12/4/2024 11:11:00 AM
Client:	Weston Solutions	Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169
Contact:	Nathan Fretz	Location:	L61

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5076-01	TAPIAL2-SB02I-7.5-1 20224-00-T1	SOIL			12/02/24 15:00			12/04/24
			pH	9045D			12/05/24 10:10	
			TOC	9060A			12/06/24 15:17	





# SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14

## Report of Analysis

Client:	Weston Solutions	Date Collected:	12/02/24 15:00
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Date Received:	12/04/24
Client Sample ID:	TAPIAL2-SB02I-7.5-120224-00-T1	SDG No.:	P5076
Lab Sample ID:	P5076-01	Matrix:	SOIL
		% Solid:	93.9

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
pH	5.27	H	1	0	0	0	pH		12/05/24 10:10	9045D
TOC	1140		1	19.8	50.0	250	mg/Kg		12/06/24 15:17	9060A

Comments: pH result reported at temperature 20.6 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



# QC RESULT SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

## Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** P5076

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB133750

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV</b> pH	pH	7.00	7	100	90-110	12/05/2024
Sample ID: <b>CCV1</b> pH	pH	2.02	2.00	101	90-110	12/05/2024
Sample ID: <b>CCV2</b> pH	pH	12.02	12.00	100	90-110	12/05/2024

## Initial and Continuing Calibration Verification

**Client:** Weston Solutions

**SDG No.:** P5076

**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

**RunNo.:** LB133779

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV1</b> TOC	mg/L	934	1000	93	90-110	11/12/2024
Sample ID: <b>CCV1</b> TOC	mg/L	937	1000	94	90-110	12/06/2024
Sample ID: <b>CCV2</b> TOC	mg/L	999	1000	100	90-110	12/06/2024

### Initial and Continuing Calibration Blank Summary

<b>Client:</b> Weston Solutions	<b>SDG No.:</b> P5076
<b>Project:</b> Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>RunNo.:</b> LB133779

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: <b>ICB1</b> TOC	mg/L	< 125.0000	125.0000	U	22.3	250	11/12/2024
Sample ID: <b>CCB1</b> TOC	mg/L	< 125.0000	125.0000	U	22.3	250	12/06/2024
Sample ID: <b>CCB2</b> TOC	mg/L	< 125.0000	125.0000	U	22.3	250	12/06/2024

**Preparation Blank Summary**

**Client:** Weston Solutions **SDG No.:** P5076  
**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: TOC	<b>LB133779BLS</b> mg/Kg	< 125.0000	125.0000	U	19.8	250	12/06/2024

## Matrix Spike Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5076
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5022-01
<b>Client ID:</b>	TAPIAL2-SB02D-13-112424-00-T1MS	<b>Percent Solids for Spike Sample:</b>	92.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	2560		1670		1000	1	89		12/06/2024



## Matrix Spike Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5076
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5022-01
<b>Client ID:</b>	TAPIAL2-SB02D-13-112424-00-T1MSD	<b>Percent Solids for Spike Sample:</b>	92.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	2550		1670		1000	1	88		12/06/2024

### Duplicate Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5076
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5022-01
<b>Client ID:</b>	TAPIAL2-SB02D-13-112424-00-T1MSD	<b>Percent Solids for Spike Sample:</b>	92.9

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TOC	mg/Kg	+/-20	2560		2550		1	0		12/06/2024

### Duplicate Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5076
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Sample ID:</b>	P5076-01
<b>Client ID:</b>	TAPIAL2-SB02I-7.5-120224-00-T1DUP	<b>Percent Solids for Spike Sample:</b>	93.9

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
pH	pH	+/-20	5.27		5.28		1	0.19		12/05/2024

### Laboratory Control Sample Summary

<b>Client:</b>	Weston Solutions	<b>SDG No.:</b>	P5076
<b>Project:</b>	Ft Meade Tipton Airfield Parcel RI - PO 0111169	<b>Run No.:</b>	LB133779

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133779BSS							
TOC	mg/Kg	1000	962		96	1	90-110	12/06/2024



# RAW DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133750

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/05/2024	08:45
2	CAL2	1	Water	NA	NA	20.2	7.01	12/05/2024	08:46
3	CAL3	1	Water	NA	NA	20.3	10.02	12/05/2024	08:48
4	ICV	1	Water	NA	NA	20.3	7.00	12/05/2024	08:50
5	CCV1	1	Water	NA	NA	20.3	2.02	12/05/2024	10:00
6	P5076-01	1	Solid	20.02	20	20.6	5.27	12/05/2024	10:10
7	P5076-01DUP	1	Solid	20.03	20	20.7	5.28	12/05/2024	10:11
8	CCV2	1	Water	NA	NA	20.3	12.02	12/05/2024	10:15

WORKLIST(Hardcopy Internal Chain)

133750

WorkList Name : ph p5076      WorkList ID : 185992      Department : Wet-Chemistry      Date : 12-05-2024 08:40:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5076-01	TAPIAL2-SB02I-7.5-120224-00-	Solid	pH	Cool 4 deg C	WEST04	L61	12/02/2024	9045D

Date/Time 12/05/24 08:47  
Raw Sample Received by: JH WDC  
Raw Sample Relinquished by: JH WDC

Date/Time 12/05/24 13:00  
Raw Sample Received by: JH WDC  
Raw Sample Relinquished by: JH WDC

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
CCV1	906.9207			TOC	
CCV1	946.1523			TOC	
CCV1	943.5790			TOC	
CCV1.....	952.4374...	..	...	TOC	..
CCB1	24.7643			TOC	
CCB1	7.8669			TOC	
CCB1.....	30.2333...	..	...	TOC	..
CCB1	23.6075			TOC	
LB133779BLS	2.3765			TOC	
LB133779BLS.....	1.9132...	..	...	TOC	..
LB133779BLS	5.4325			TOC	
LB133779BLS	13.4427			TOC	
LB133779BSS.....	960.5422...	..	...	TOC	..
LB133779BSS	966.7218			TOC	
LB133779BSS	945.8391			TOC	
LB133779BSS.....	973.9248...	..	...	TOC	..
P5113-01	313.2646			TOC	
P5113-01	331.6778			TOC	
P5113-01.....	232.7811...	..	...	TOC	..
P5113-01	231.7867			TOC	
P5113-02	424.3524			TOC	
P5113-02.....	348.4884...	..	...	TOC	..
P5113-02	404.6520			TOC	
P5113-02	409.2955			TOC	
P5022-01.....	1592.4867...	..	...	TOC	..
P5022-01	1520.0836			TOC	
P5022-01	1924.7922			TOC	
P5022-01.....	1631.9971...	..	...	TOC	..
P5022-01MS	2194.9075			TOC	
P5022-01MS	2786.8528			TOC	
P5022-01MS.....	2944.4070...	..	...	TOC	..
P5022-01MS	2298.9773			TOC	
P5022-01MSD	2434.9634			TOC	
P5022-01MSD.....	2164.7324...	..	...	TOC	..
P5022-01MSD	3038.4358			TOC	
P5022-01MSD	2567.5728			TOC	
P5076-01.....	1049.1886...	..	...	TOC	..
P5076-01	1247.4980			TOC	
P5076-01	1138.2893			TOC	
P5076-01.....	1132.0439...	..	...	TOC	..
P5117-01	477.3856			TOC	
P5117-01	409.1736			TOC	
P5117-01.....	378.8983...	..	...	TOC	..
P5117-01	338.3237			TOC	
CCV2	1074.5073			TOC	
CCV2.....	980.9690...	..	...	TOC	..
CCV2	974.1614			TOC	
CCV2	965.5085			TOC	
CCB2.....	5.6898...	..	...	TOC	..
CCB2	17.7748			TOC	
CCB2	43.7854			TOC	
CCB2.....	7.1315...	..	...	TOC	..

1

2

3

4

5

6

7

8

9

10

11

12

13

14



Method ID	Sample Type	Vial	Timestamp	Message
=====	=====	=====	=====	=====
Boat Sampler	Sample		2024/12/06 10:01	
Boat Sampler	Sample		2024/12/06 10:09	
Boat Sampler	Sample		2024/12/06 10:12	
Boat Sampler	...Sample	..	..2024/12/06 10:16	..
Boat Sampler	Sample		2024/12/06 10:21	
Boat Sampler	Sample		2024/12/06 10:24	Low Sample Detected
Boat Sampler	...Sample	..	..2024/12/06 10:26	..
Boat Sampler	Sample		2024/12/06 10:28	
Boat Sampler	Sample		2024/12/06 10:32	Low Sample Detected
Boat Sampler	...Sample	..	..2024/12/06 10:36	..Low Sample Detected
Boat Sampler	Sample		2024/12/06 10:52	Low Sample Detected
Boat Sampler	Sample		2024/12/06 10:59	
Boat Sampler	...Sample	..	..2024/12/06 11:04	..
Boat Sampler	Sample		2024/12/06 11:06	
Boat Sampler	Sample		2024/12/06 11:10	
Boat Sampler	...Sample	..	..2024/12/06 11:15	..
Boat Sampler	Sample		2024/12/06 11:45	
Boat Sampler	Sample		2024/12/06 11:53	
Boat Sampler	...Sample	..	..2024/12/06 12:05	..
Boat Sampler	Sample		2024/12/06 12:09	
Boat Sampler	Sample		2024/12/06 12:43	
Boat Sampler	...Sample	..	..2024/12/06 12:51	..
Boat Sampler	Sample		2024/12/06 12:55	
Boat Sampler	Sample		2024/12/06 13:00	
Boat Sampler	...Sample	..	..2024/12/06 13:17	..
Boat Sampler	Sample		2024/12/06 13:23	
Boat Sampler	Sample		2024/12/06 13:27	
Boat Sampler	...Sample	..	..2024/12/06 13:55	..
Boat Sampler	Sample		2024/12/06 14:02	
Boat Sampler	Sample		2024/12/06 14:05	
Boat Sampler	...Sample	..	..2024/12/06 14:09	..
Boat Sampler	Sample		2024/12/06 14:12	
Boat Sampler	Sample		2024/12/06 14:16	
Boat Sampler	...Sample	..	..2024/12/06 14:20	..
Boat Sampler	Sample		2024/12/06 14:23	
Boat Sampler	Sample		2024/12/06 14:25	
Boat Sampler	...Sample	..	..2024/12/06 14:51	..
Boat Sampler	Sample		2024/12/06 14:55	
Boat Sampler	Sample		2024/12/06 15:12	
Boat Sampler	...Sample	..	..2024/12/06 15:17	..
Boat Sampler	Sample		2024/12/06 15:24	
Boat Sampler	Sample		2024/12/06 15:27	
Boat Sampler	...Sample	..	..2024/12/06 15:30	..
Boat Sampler	Sample		2024/12/06 15:34	
Boat Sampler	Sample		2024/12/06 15:38	
Boat Sampler	...Sample	..	..2024/12/06 15:41	..
Boat Sampler	Sample		2024/12/06 15:43	
Boat Sampler	Sample		2024/12/06 15:46	
Boat Sampler	...Sample	..	..2024/12/06 15:50	..Low Sample Detected
Boat Sampler	Sample		2024/12/06 15:55	
Boat Sampler	Sample		2024/12/06 15:59	
Boat Sampler	...Sample	..	..2024/12/06 16:02	..Low Sample Detected

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Sample ID: CCB1 Mode: TOC  
Method: Boat Sampler Filename: 12061025  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 10:26  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	30.2333	1.2093	85634	-2.969	-1.983	37

Sample ID: CCB1 Mode: TOC  
Method: Boat Sampler Filename: 12061027  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 10:28  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	23.6075	0.9443	66867	-2.937	-1.944	33

Sample ID: LB133779BLS Mode: TOC  
Method: Boat Sampler Filename: 12061029  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 10:32  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2.3765	0.0951	6731	-2.943	-3.029	120

Last Message: Low Sample Detected

Sample ID: LB133779BLS Mode: TOC  
Method: Boat Sampler Filename: 12061033  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 10:36  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1.9132	0.0765	5419	-2.979	-3.048	120

Last Message: Low Sample Detected

Sample ID: LB133779BLS Mode: TOC  
Method: Boat Sampler Filename: 12061049  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 10:52  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	5.4325	0.2173	15387	-3.102	-3.151	120

Last Message: Low Sample Detected

Sample ID: LB133779BLS Mode: TOC  
Method: Boat Sampler Filename: 12061058  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 10:59  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	13.4427	0.5377	38076	-3.204	-2.207	32

Sample ID: LB133779BSS Mode: TOC  
Method: Boat Sampler Filename: 12061102  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 11:04  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	960.5422	38.4217	2720693	-3.023	-2.028	65

Sample ID: LB133779BSS Mode: TOC  
Method: Boat Sampler Filename: 12061104  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 11:06  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	966.7218	38.6689	2738197	-2.960	-1.965	67

Sample ID: LB133779BSS Mode: TOC  
Method: Boat Sampler Filename: 12061108  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 11:10  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	945.8391	37.8336	2679048	-3.196	-2.200	68

Sample ID: LB133779BSS Mode: TOC  
Method: Boat Sampler Filename: 12061113  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 11:15  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	973.9248	38.9570	2758599	-3.246	-2.248	69

Sample ID: P5113-01 Mode: TOC  
Method: Boat Sampler Filename: 12061143  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 11:45  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	313.2646	1.9422	137533	-3.371	-2.381	36

Sample ID: P5113-01 Mode: TOC  
Method: Boat Sampler Filename: 12061152  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 11:53  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	331.6778	2.4212	171452	-3.383	-2.390	37

Sample ID: P5113-01  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061204  
Timestamp: 2024/12/06 12:05  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	232.7811	1.6062	113737	-3.366	-2.367	36

Sample ID: P5113-01  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061208  
Timestamp: 2024/12/06 12:09  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	231.7867	1.5761	111609	-3.419	-2.425	34

Sample ID: P5113-02  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061242  
Timestamp: 2024/12/06 12:43  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	424.3524	3.0553	216353	-3.366	-2.373	42

Sample ID: P5113-02  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061250  
Timestamp: 2024/12/06 12:51  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	348.4884	2.0909	148062	-3.386	-2.387	42

Sample ID: P5113-02  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061254  
Timestamp: 2024/12/06 12:55  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	404.6520	2.2661	160462	-3.350	-2.358	40

Sample ID: P5113-02  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061259  
Timestamp: 2024/12/06 13:00  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
-------	-------	------	----------	-----------------------	--------------------	---------------------

1	409.2955	2.4558	173897	-3.416	-2.418	41
=====						

Sample ID:	P5022-01	Mode:	TOC
Method:	Boat Sampler	Filename:	12061315
Cal. Curve:	TOC SOIL	Timestamp:	2024/12/06 13:17
Operator ID:	NF IZ	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1592.4867	12.7399	902130	-3.460	-2.463	53
=====						

Sample ID:	P5022-01	Mode:	TOC
Method:	Boat Sampler	Filename:	12061322
Cal. Curve:	TOC SOIL	Timestamp:	2024/12/06 13:23
Operator ID:	NF IZ	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1520.0836	10.3366	731947	-3.406	-2.412	51
=====						

Sample ID:	P5022-01	Mode:	TOC
Method:	Boat Sampler	Filename:	12061325
Cal. Curve:	TOC SOIL	Timestamp:	2024/12/06 13:27
Operator ID:	NF IZ	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1924.7922	11.1638	790524	-3.446	-2.453	52
=====						

Sample ID:	P5022-01	Mode:	TOC
Method:	Boat Sampler	Filename:	12061354
Cal. Curve:	TOC SOIL	Timestamp:	2024/12/06 13:55
Operator ID:	NF IZ	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1631.9971	10.2816	728053	-3.216	-2.222	49
=====						

Sample ID:	P5022-01MS	Mode:	TOC
Method:	Boat Sampler	Filename:	12061400
Cal. Curve:	TOC SOIL	Timestamp:	2024/12/06 14:02
Operator ID:	NF IZ	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2194.9075	12.7305	901462	-3.349	-2.353	54
=====						

Sample ID:	P5022-01MS	Mode:	TOC
Method:	Boat Sampler	Filename:	12061403
Cal. Curve:	TOC SOIL	Timestamp:	2024/12/06 14:05
Operator ID:	NF IZ	Sample Type:	Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2786.8528	15.3277	1085375	-3.316	-2.321	59
=====						

Sample ID: P5022-01MS Mode: TOC  
Method: Boat Sampler Filename: 12061407  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:09  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2944.4070	18.2553	1292685	-3.368	-2.374	62

Sample ID: P5022-01MS Mode: TOC  
Method: Boat Sampler Filename: 12061411  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:12  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2298.9773	14.2537	1009322	-3.389	-2.396	57

Sample ID: P5022-01MSD Mode: TOC  
Method: Boat Sampler Filename: 12061414  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:16  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2434.9634	16.5578	1172477	-3.359	-2.360	59

Sample ID: P5022-01MSD Mode: TOC  
Method: Boat Sampler Filename: 12061418  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:20  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2164.7324	12.9884	919726	-3.330	-2.331	54

Sample ID: P5022-01MSD Mode: TOC  
Method: Boat Sampler Filename: 12061421  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:23  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	3038.4358	18.8383	1333967	-3.342	-2.344	61

Sample ID: P5022-01MSD Mode: TOC  
Method: Boat Sampler Filename: 12061424  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:25  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2567.5728	14.8919	1054518	-3.298	-2.298	57

Sample ID: P5076-01 Mode: TOC  
Method: Boat Sampler Filename: 12061449  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:51  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1049.1886	5.6656	401190	-3.439	-2.439	46

Sample ID: P5076-01 Mode: TOC  
Method: Boat Sampler Filename: 12061453  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 14:55  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1247.4980	6.7365	477020	-3.456	-2.459	47

Sample ID: P5076-01 Mode: TOC  
Method: Boat Sampler Filename: 12061511  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 15:12  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1138.2893	6.6021	467502	-3.417	-2.425	50

Sample ID: P5076-01 Mode: TOC  
Method: Boat Sampler Filename: 12061516  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 15:17  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1132.0439	6.2262	440889	-3.397	-2.398	45

Sample ID: P5117-01 Mode: TOC  
Method: Boat Sampler Filename: 12061522  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 15:24  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	477.3856	3.4372	243391	-3.413	-2.424	39

Sample ID: P5117-01 Mode: TOC  
Method: Boat Sampler Filename: 12061526  
Cal. Curve: TOC SOIL Timestamp: 2024/12/06 15:27  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	409.1736	2.7824	197024	-3.384	-2.392	38

Sample ID: P5117-01 Mode: TOC



Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Filename: 12061529  
Timestamp: 2024/12/06 15:30  
Sample Type: Sample

Reviewed By:Iwona  
On:12/10/2024 9:22:06  
AM  
Inst Id :Appolo-9000  
LB :LB133779

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	378.8983	3.0312	214643	-3.379	-2.382	36
=====						

Sample ID: P5117-01  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061533  
Timestamp: 2024/12/06 15:34  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	338.3237	1.9623	138952	-3.396	-2.402	36
=====						

Sample ID: CCV2  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061535  
Timestamp: 2024/12/06 15:38  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	1074.5073	42.9803	3043495	-3.365	-2.366	88
=====						

Sample ID: CCV2  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061539  
Timestamp: 2024/12/06 15:41  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	980.9690	39.2388	2778551	-3.194	-2.195	71
=====						

Sample ID: CCV2  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061542  
Timestamp: 2024/12/06 15:43  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	974.1614	38.9665	2759269	-3.147	-2.151	71
=====						

Sample ID: CCV2  
Method: Boat Sampler  
Cal. Curve: TOC SOIL  
Operator ID: NF IZ

Mode: TOC  
Filename: 12061544  
Timestamp: 2024/12/06 15:46  
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	965.5085	38.6203	2734760	-3.083	-2.083	68
=====						

Sample ID: CCB2  
Method: Boat Sampler  
Cal. Curve: TOC SOIL

Mode: TOC  
Filename: 12061547  
Timestamp: 2024/12/06 15:50

Operator ID: NF IZ				Sample Type: Sample			Reviewed By: Iwona On: 12/10/2024 9:22:06 AM Inst Id : Appolo-9000 LB : LB133779
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time	
1	5.6898	0.2276	16116	-3.126	-3.264	120	
-----							
Last Message: Low Sample Detected							
=====							
Sample ID: CCB2				Mode:	TOC		
Method: Boat Sampler				Filename:	12061554		
Cal. Curve: TOC SOIL							
Timestamp: 2024/12/06 15:55							
Operator ID: NF IZ				Sample Type: Sample			
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time	
1	17.7748	0.7110	50346	-3.273	-2.286	32	
=====							
Sample ID: CCB2				Mode:	TOC		
Method: Boat Sampler				Filename:	12061558		
Cal. Curve: TOC SOIL				Timestamp:	2024/12/06 15:59		
Operator ID: NF IZ				Sample Type: Sample			
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time	
1	43.7854	1.7514	124020	-3.256	-2.258	42	
=====							
Sample ID: CCB2				Mode:	TOC		
Method: Boat Sampler				Filename:	12061559		
Cal. Curve: TOC SOIL				Timestamp:	2024/12/06 16:02		
Operator ID: NF IZ				Sample Type: Sample			
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time	
1	7.1315	0.2853	20200	-3.217	-3.281	120	
-----							
Last Message: Low Sample Detected							
=====							

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
=====	=====	=====	=====	=====	=====
BLANK	9263	5695	61.48	TOC	
250mg/l	813930	31881	3.92	TOC	
500mg/l	1580916	138813	8.78	TOC	
1000mg/l.....	2797885...	74757..	2.67...	TOC	..
2000mg/l	5752648	21216	0.37	TOC	
ICV	943.9813			TOC	
ICV.....	919.2893...	..	...	TOC	..
ICV	940.0988			TOC	
ICV	933.3373			TOC	
ICB.....	4.4514...	..	...	TOC	..
ICB	9.3085			TOC	
ICB	6.7305			TOC	
ICB.....	7.5078...	..	...	TOC	..

1

2

3

4

5

6

7

8

9

10

11

12

13

14

Method ID	Sample Type	Vial Timestamp		Message
=====				
Boat Sampler	TOC Standard	2024/11/12	11:01	Low Sample Detected
Boat Sampler	TOC Standard	2024/11/12	11:15	
Boat Sampler	TOC Standard	2024/11/12	11:27	
Boat Sampler	...TOC Standard	..	..2024/11/12 11:36	..
Boat Sampler	TOC Standard	2024/11/12	12:05	
Boat Sampler	Sample	2024/11/12	12:12	
Boat Sampler	...Sample	..	..2024/11/12 12:14	..
Boat Sampler	Sample	2024/11/12	12:16	
Boat Sampler	Sample	2024/11/12	12:18	
Boat Sampler	...Sample	..	..2024/11/12 12:23	..Low Sample Detected
Boat Sampler	Sample	2024/11/12	12:26	Low Sample Detected
Boat Sampler	Sample	2024/11/12	12:29	Low Sample Detected
Boat Sampler	...Sample	..	..2024/11/12 12:32	..Low Sample Detected

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Sample ID: 2000mg/l Mode: TOC  
Method: Boat Sampler Filename: 11121155  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:05  
Operator ID: NF IZ Sample Type: TOC Standard

Reviewed By:Iwona  
On:12/10/2024 9:22:06  
AM  
Inst Id :Appolo-9000  
LB :LB133779

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			5772363	-2.492	-1.493	116
2			5763448	-2.138	-1.139	91
3			5723617	-0.015	0.979	57
4			5751163	-0.355	0.644	62

<<<Statistics>>> Mean: 5752648 Std Dev: 21216 RSD: 0.37

Sample ID: ICV Mode: TOC  
Method: Boat Sampler Filename: 11121210  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:12  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	943.9813	37.7593	2673786	-2.569	-1.574	83

Sample ID: ICV Mode: TOC  
Method: Boat Sampler Filename: 11121212  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:14  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	919.2893	36.7716	2603846	-2.379	-1.380	64

Sample ID: ICV Mode: TOC  
Method: Boat Sampler Filename: 11121214  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:16  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	940.0988	37.6040	2662788	-2.346	-1.346	61

Sample ID: ICV Mode: TOC  
Method: Boat Sampler Filename: 11121217  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:18  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	933.3373	37.3335	2643637	-2.443	-1.445	63

Sample ID: ICB Mode: TOC  
Method: Boat Sampler Filename: 11121220  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:23  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	4.4514	0.1781	12608	-2.598	-2.685	120

-----  
Last Message: Low Sample Detected  
=====

Sample ID: ICB Mode: TOC  
Method: Boat Sampler Filename: 11121223  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:26  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	9.3085	0.3723	26366	-2.661	-2.693	120

-----

Last Message: Low Sample Detected  
=====

Sample ID: ICB Mode: TOC  
Method: Boat Sampler Filename: 11121226  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:29  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	6.7305	0.2692	19064	-2.649	-2.682	120

-----

Last Message: Low Sample Detected  
=====

Sample ID: ICB Mode: TOC  
Method: Boat Sampler Filename: 11121230  
Cal. Curve: TOC SOIL Timestamp: 2024/11/12 12:32  
Operator ID: NF IZ Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	7.5078	0.3003	2			
1266	-2.657	-2.679	120			

-----

Last Message: Low Sample Detected  
=====

Calibration Report    Print Date/Time: 2024/11/12 12:05:31

Cal. Curve ID:                TOC SOIL  
Created:                      2024/11/12 12:05  
Calibration Factor (m): 7.081e+04  
Y Intercept (b):            66586  
r-squared:                   0.99875

Standard ID	Y	X Expected	Measured	Re Message	Date & Time
	Raw Data	ug C	ug C		
BLANK	9263	0.000	-0.810		2024/11/12 11:01
250mg/l	813930	10.000	10.554	5.5	2024/11/12 11:15
500mg/l	1580917	20.000	21.385	6.9	2024/11/12 11:27
1000mg/l	2797884	40.000	38.571	-3.6	2024/11/12 11:36
2000mg/l	5752648	80.000	80.299	0.4	2024/11/12 12:05

12  
11/12/24



WORKLIST(Hardcopy Internal Chain)

LB133779

WorkList Name : TOC SOIL-12032024

WorkList ID : 185934

Department : Wet-Chemistry

Date : 12-03-2024 15:04:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5113-01	FES-SB406-4345	Solid	TOC	Cool 4 deg C	TETR06	L31	12/04/2024	9060A
P5113-02	FES-SB406-7375	Solid	TOC	Cool 4 deg C	TETR06	L31	12/04/2024	9060A
P5022-01	TAPIAL2-SB02D-13-112424-00-	Solid	TOC	Cool 4 deg C	WEST04	L41	11/24/2024	9060A
P5076-01	TAPIAL2-SB02I-7.5-120224-00-	Solid	TOC	Cool 4 deg C	WEST04	L61	12/02/2024	9060A
P5117-01	TAPIAL3-SB04I-10-120324-00-	Solid	TOC	Cool 4 deg C	WEST04	L41	12/05/2024	9060A

Date/Time 12.06.2024, 08:30  
Raw Sample Received by: NFWC  
Raw Sample Relinquished by: NFWC

Date/Time 12.06.2024, 11:00  
Raw Sample Received by: NFWC  
Raw Sample Relinquished by: NFWC

**Instrument ID:** WC PH METER-1

**Daily Analysis Runlog For Sequence/QC Batch ID # LB133750**

Review By	jignesh	Review On	12/5/2024 9:05:52 AM
Supervise By	Iwona	Supervise On	12/5/2024 9:46:10 AM
SubDirectory	LB133750	Test	pH
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3107,W3093,W3094,W3071,W3005,W3072		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	12/05/24 08:45		Jignesh	OK
2	CAL2	CAL2	CAL	12/05/24 08:46		Jignesh	OK
3	CAL3	CAL3	CAL	12/05/24 08:48		Jignesh	OK
4	ICV	ICV	ICV	12/05/24 08:50		Jignesh	OK
5	CCV1	CCV1	CCV	12/05/24 10:00		Jignesh	OK
6	P5076-01	TAPIAL2-SB02I-7.5-1	SAM	12/05/24 10:10		Jignesh	OK
7	P5076-01DUP	TAPIAL2-SB02I-7.5-1	DUP	12/05/24 10:11		Jignesh	OK
8	CCV2	CCV2	CCV	12/05/24 10:15		Jignesh	OK

Instrument ID: TOC

**Daily Analysis Runlog For Sequence/QC Batch ID # LB133779**

Review By	Niha	Review On	12/9/2024 2:36:05 PM
Supervise By	Iwona	Supervise On	12/10/2024 9:22:06 AM
SubDirectory	LB133779	Test	TOC
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	WP110667,WP110662,WP110663,WP110664,WP110665		
ICV Standard	WP110666		
CCV Standard	WP111002		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111003		
Chk Standard	WP109225		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	BLANK	BLANK	CAL1	11/12/24 11:01		NF IZ	OK
2	250mg/l	250mg/l	CAL2	11/12/24 11:15		NF IZ	OK
3	500mg/l	500mg/l	CAL3	11/12/24 11:27		NF IZ	OK
4	1000mg/l	1000mg/l	CAL4	11/12/24 11:36		NF IZ	OK
5	2000mg/l	2000mg/l	CAL5	11/12/24 12:05		NF IZ	OK
6	ICV1	ICV1	ICV	11/12/24 12:18		NF IZ	OK
7	ICB1	ICB1	ICB	11/12/24 12:32		NF IZ	OK
8	CCV1	CCV1	CCV	12/06/24 10:16		NF IZ	OK
9	CCB1	CCB1	CCB	12/06/24 10:28		NF IZ	OK
10	LB133779BLS	LB133779BLS	MB	12/06/24 10:59		NF IZ	OK
11	LB133779BSS	LB133779BSS	LCS	12/06/24 11:15		NF IZ	OK
12	P5113-01	FES-SB406-4345	SAM	12/06/24 12:09		NF IZ	OK
13	P5113-02	FES-SB406-7375	SAM	12/06/24 13:00		NF IZ	OK
14	P5022-01	TAPIAL2-SB02D-13-1	SAM	12/06/24 13:55		NF IZ	OK
15	P5022-01MS	TAPIAL2-SB02D-13-1	MS	12/06/24 14:12	sample + 40ul of 111002	NF IZ	OK
16	P5022-01MSD	TAPIAL2-SB02D-13-1	MSD	12/06/24 14:25	sample + 40ul of 111002	NF IZ	OK
17	P5076-01	TAPIAL2-SB02I-7.5-1	SAM	12/06/24 15:17		NF IZ	OK
18	P5117-01	TAPIAL3-SB04I-10-12	SAM	12/06/24 15:34		NF IZ	OK

Instrument ID: TOC

**Daily Analysis Runlog For Sequence/QC Batch ID # LB133779**

Review By	Niha	Review On	12/9/2024 2:36:05 PM
Supervise By	Iwona	Supervise On	12/10/2024 9:22:06 AM
SubDirectory	LB133779	Test	TOC

STD. NAME	STD REF.#
ICAL Standard	WP110667,WP110662,WP110663,WP110664,WP110665
ICV Standard	WP110666
CCV Standard	WP111002
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	WP111003
Chk Standard	WP109225

19	CCV2	CCV2	CCV	12/06/24 15:46		NF IZ	OK
20	CCB2	CCB2	CCB	12/06/24 16:02		NF IZ	OK

## Prep Standard - Chemical Standard Summary

**Order ID :** P5076

**Test :** Percent Solids,pH,TOC

**Prepbatch ID :**

**Sequence ID/Qc Batch ID:** LB133750,LB133779,

**Standard ID :**

WP109217,WP109218,WP109225,WP110662,WP110663,WP110664,WP110665,WP110666,WP110667,WP111002,W  
P111003,

**Chemical ID :**

W2784,W2860,W3005,W3071,W3072,W3093,W3094,W3107,W3111,W3112,

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2050	TOC STOCK STD, 4000PPM	<a href="#">WP109217</a>	08/07/2024	01/18/2025	Iwona Zarych	WETCHEM_SCALE_5 (WC)	WETCHEM_FIPETTE_3 (WC)	Mohan Bera 08/16/2024
<b>FROM</b> 5.00000ml of W2860 + 8.51200gram of W3111 + 990.00000ml of W3112 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2051	TOC STOCK STD-SS, 4000PPM	<a href="#">WP109218</a>	08/07/2024	02/07/2025	Iwona Zarych	WETCHEM_SCALE_5 (WC)	WETCHEM_FIPETTE_3 (WC)	Mohan Bera 08/16/2024
<b>FROM</b> 5.00000ml of W2860 + 8.51200gram of W2784 + 990.00000ml of W3112 = Final Quantity: 1000.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2435	1:1 PHOSPHORIC ACID FOR TOC SOILS	<a href="#">WP109225</a>	08/07/2024	02/07/2025	Iwona Zarych	None	WETCHEM_FIPETTE_3 (WC)	Mohan Bera 08/16/2024
<b>FROM</b> 50.00000ml of W2860 + 50.00000ml of W3112 = Final Quantity: 100.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
712	TOC SOIL cal 250ppm	<a href="#">WP110662</a>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	Iwona Zarych 11/14/2024
<b>FROM</b> 15.00000ml of W3112 + 1.00000ml of WP109217 = Final Quantity: 16.000 ml								

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
710	TOC SOIL cal 500ppm	<a href="#">WP110663</a>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	Iwona Zarych 11/14/2024

**FROM** 14.00000ml of W3112 + 2.00000ml of WP109217 = Final Quantity: 16.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3544	TOC SOIL Cal- CCV 1000PPM	<a href="#">WP110664</a>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	Iwona Zarych 11/14/2024

**FROM** 15.00000ml of W3112 + 5.00000ml of WP109217 = Final Quantity: 20.000 ml



## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
713	TOC SOIL cal 2000ppm	<a href="#">WP110665</a>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	Iwona Zarych 11/14/2024

**FROM** 5.00000ml of W3112 + 5.00000ml of WP109217 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2819	TOC ICV-LCSS, 1000PPM	<a href="#">WP110666</a>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	Iwona Zarych 11/14/2024

**FROM** 15.00000ml of W3112 + 5.00000ml of WP109218 = Final Quantity: 20.000 ml

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
304	TOC CAL 0.00ppm	<a href="#">WP110667</a>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	Iwona Zarych 11/14/2024

**FROM** 100.00000ml of W3112 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3544	TOC SOIL Cal- CCV 1000PPM	<a href="#">WP111002</a>	12/06/2024	12/13/2024	Niha Farheen Shaik	None	None	Iwona Zarych 12/09/2024

**FROM** 15.00000ml of W3112 + 5.00000ml of WP109217 = Final Quantity: 20.000 ml

## Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2819	TOC ICV-LCSS, 1000PPM	<a href="#">WP111003</a>	12/06/2024	12/13/2024	Niha Farheen Shaik	None	None	Iwona Zarych 12/09/2024
<b>FROM</b> 15.00000ml of W3112 + 5.00000ml of WP109218 = Final Quantity: 20.000 ml								

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	P243-500 / Potassium Hydrogen Phthalate, 500 gms	201089	06/30/2025	12/23/2020 / apatel	12/16/2020 / apatel	W2784

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J0260-3 / Phosphoric Acid, 2.5 L	0000278313	01/31/2026	07/12/2021 / apatel	07/12/2021 / apatel	W2860

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL13850-1 / Buffer Solution, PH2 (500ml)	4212E45	12/31/2024	01/31/2023 / lwona	01/31/2023 / lwona	W3005

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14455-3 / buffer solution pH 7 yellow	4308H30	07/31/2025	01/02/2024 / JIGNESH	12/06/2023 / lwona	W3071

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14940-1 / Buffer Solution, PH12 (500ml)	2310P21	04/30/2025	01/02/2024 / JIGNESH	12/07/2023 / lwona	W3072

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	566002 / BUFFER PH 7.00 GREEN 1PINT PK6	44001f99	12/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3093

### CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	1601-1 / PH 10.01 BUFFER,COLOR CD 475ML	4310g83	03/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3094

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14055-3 / PH 4 BUFFER SOLUTION	AL14055-3	02/27/2026	09/05/2024 / jignesh	05/13/2024 / jignesh	W3107

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	P243-500 / Potassium Hydrogen Phthalate, 500 gms	24A1956910	01/18/2025	06/26/2024 / lwona	06/26/2024 / lwona	W3111

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Phosphoric Acid  
BAKER ANALYZED® A.C.S. Reagent

(orthophosphoric acid)



Material No.: 0260-03  
Batch No.: 0000278313  
Manufactured Date: 2021/02/01  
Retest Date: 2026/01/31  
Revision No: 2

## Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay (H <sub>3</sub> PO <sub>4</sub> ) (by acidimetry)	85.0 – 87.0 %	85.8
Calcium (Ca)	<= 0.002 %	< 0.001
Color (APHA)	<= 10	5
Insoluble Matter	<= 0.001 %	< 0.001
ACS – Magnesium (Mg)	<= 0.002 %	<0.002
Sulfate (SO <sub>4</sub> )	<= 12 ppm	< 4
Volatile Acids (as CH <sub>3</sub> COOH)	<= 0.001 %	0.001
Reducing Substances	Passes Test	PT
Chloride (Cl)	<= 3 ppm	< 1
Nitrate (NO <sub>3</sub> )	<= 5 ppm	< 2
Trace Impurities – Antimony (Sb)	<= 20.000 ppm	0.007
Trace Impurities – Arsenic (As)	<= 0.500 ppm	< 0.001
Trace Impurities – Iron (Fe)	<= 10.000 ppm	< 1.000
Heavy Metals (as Pb)	<= 8 ppm	< 3
Trace Impurities – Manganese (Mn)	<= 0.500 ppm	0.005
Trace Impurities – Potassium (K)	<= 40.000 ppm	< 0.001
Trace Impurities – Sodium (Na)	<= 200.000 ppm	0.082

For Laboratory, Research or Manufacturing Use  
Exceeds A.C.S. Specifications  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US  
Packaging Site: Phillipsburg Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

W3071  
Rec 12/6/23

## Certificate of Analysis 12

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4308H30

Product Number: 1551

Manufacture Date: AUG 09, 2023

Expiration Date: JUL 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Yellow liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.002	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (08/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.





# RICCA CHEMICAL COMPANY®

1841 Broad Street  
Pocomoke City, MD 21851  
<http://www.riccachemical.com>  
1-888-GO-RICCA  
[customerservice@riccachemical.com](mailto:customerservice@riccachemical.com)

W 3072  
REC. 12/01/23  
12

## Certificate of Analysis

Buffer, Reference Standard, pH 12.00 ± 0.01 at 25°C

Lot Number: 2310P21

Product Number: 1615

Manufacture Date: OCT 24, 2023

Expiration Date: APR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C	15	20	25	30	35	40
pH	12.35	12.17	11.99	11.78	11.62	11.46

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	12.005	0.02	186-I-g, 186-II-g, 191d

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1615-1	4 L natural poly	18 months
1615-16	500 mL clear PET-G	18 months
1615-32	1 L natural poly	18 months
1615-5	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

*Sharon Travers*

Sharon Travers (10/24/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC

## Certificate of Analysis

1 Reagent Lane  
Fair Lawn, NJ 07410  
201.796.7100 tel  
201.796.1329 faxThermo Fisher Scientific's Quality System has been found to conform to Quality Management System  
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	P243	Quality Test / Release Date	06/19/2020
Lot Number	201089		
Description	POTASSIUM HYDROGEN PHTHALATE, ACIDIMETRIC STANDARD, A.C.S.		
Country of Origin	Spain	Suggested Retest Date	Jun/2025
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	WHITE CRYSTALS
ASSAY POTASSIUM HYDROGEN PHTHALATE	%	Inclusive Between 99.95 - 100.05	100.03
CHLORINE COMPOUNDS	%	<= 0.003	<0.003
HEAVY METALS (as Pb)	ppm	<= 5	<5
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
INSOLUBLE MATTER	%	<= 0.005	<0.005
IRON (Fe)	ppm	<= 5	<5
PH OF 0.05M SOLUTION		Inclusive Between 4.00 - 4.02	4.00
SODIUM (Na)	%	<= 0.005	<0.005
SULFUR COMPOUNDS	%	<= 0.002	<0.002%
TRACEABLE TO NIST	SOD CARBONATE	= LOT 351a	351a
TRACEABLE TO NIST KHP STD	POT. ACID PHTHALATE	= LOT 84L	84L



Julian Burton - Quality Control Manager – Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.  
If there are any questions with this certificate, please call at (800) 227-6701.

\*Based on suggested storage condition.



# RICCA CHEMICAL COMPANY®

W 3005

REC- 1/31/23

12

1490 Lammers Pike

Batesville, IN 47006

<http://www.riccachemical.com>

1-888-GO-RICCA

customerservice@riccachemical.com

## Certificate of Analysis

Buffer, Reference Standard, pH 2.00 ± 0.01 at 25°C

Lot Number: 4212E45

Product Number: 1493

Manufacture Date: DEC 20, 2022

Expiration Date: DEC 2024

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	10	15	20	25	30	35	40	45	50
pH	1.93	1.98	1.98	2.00	2.01	2.03	2.03	2.04	2.04

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	2.000	0.02	185i, 186-I-g, 186-II-g

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1493-1	4 L natural poly	24 months
1493-16	500 mL natural poly	24 months
1493-32	1 L natural poly	24 months
1493-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (12/20/2022)

Production Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



## Certificate of Analysis

W3093  
094121  
04/03/2024  
16

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4401F99

Product Number: 1551

Manufacture Date: JAN 08, 2024

Expiration Date: DEC 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	
Sodium Hydroxide	1310-73-2	

Test	Specification	Result
Appearance	Yellow liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.004	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-1	4 L natural poly	24 months
1551-1CT	4 L Cubitainer®	24 months
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

*Paul Brandon*

Paul Brandon (01/08/2024)  
Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.





## Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 4310G83

Product Number: 1601

Manufacture Date: OCT 09, 2023

Expiration Date: MAR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	50
pH	10.31	10.23	10.17	10.11	10.05	10.00	9.95	9.91	9.87	9.81

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Carbonate	497-19-8	ACS
Sodium Bicarbonate	144-55-8	ACS
Sodium Hydroxide	1310-73-2	Reagent
Preservative	Proprietary	
Blue Dye	Proprietary	

Test	Specification	Result
Appearance	Blue liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	10.003	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer C	ASTM (D 5464)
Buffer C	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1601-16	500 mL natural poly	18 months
1601-5	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)





Paul Brandon (10/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



## Certificate of Analysis

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 4403F90

Product Number: 1501

Manufacture Date: MAR 09, 2024

Expiration Date: FEB 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	4.00	4.00	4.00	4.00	4.00	4.00	4.01	4.02	4.03	4.04	4.06

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	Buffer
Preservative	Proprietary	Commercial
Red Dye	Proprietary	Purified

Test	Specification	Result
Appearance	Red liquid	Passed

\*Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	4.000	0.02	185i, 186-I-g, 186-II-g

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer B	ASTM (D 5464)
Buffer B	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1501-2.5	10 L Cubitainer®	24 months
1501-32	1 L natural poly	24 months
1501-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/09/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



## POTASSIUM HYDROGEN PHTHALATE

**Material:** N983  
**Grade:** ACS GRADE  
**Batch Number:** 24A1956910

Chemical Formula: HOCC6H4COOK  
Molecular Weight: 204.22  
CAS #: 877-24-7  
Appearance:

Manufacture Date: 01/19/2022  
Reassay Date: 01/18/2025

Storage: Room Temperature

White crystals.

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Assay (dried basis)	99.95 - 100.05 %	99.97 %	PASS
Chlorine Compounds	<= 0.003 %	<0.003 %	PASS
Heavy Metals (as Pb)	<= 5 ppm	<5 ppm	PASS
Insoluble Matter	<= 0.005 %	0.003 %	PASS
Iron	<= 5 ppm	<5 ppm	PASS
pH (0.05M, Water) @25C	4.00 - 4.02	4.00	PASS
Sodium	<= 0.005 %	<0.005 %	PASS
Sulfur Compounds	<= 0.002 %	<0.002 %	PASS

Spec Set: N983ACS

Internal ID #: 710

Signature	Additional Information
<p>We certify that this batch conforms to the specifications listed.</p> <p>This document has been electronically produced and is valid without a signature.</p> <p>Leona Edwardson, Quality Control Sr. Manager - Solon  VWR Chemicals, LLC.  28600 Fountain Parkway, Solon OH 44139 USA</p>	<p>Analysis may have been rounded to significant digits in specification limits.</p> <p>Product meets analytical specifications of the grades listed.</p>

# PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/5/2024

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:30  
In Date: 12/04/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:14  
Out Date: 12/05/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB133737

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P5076-01	TAPIAL2-SB02I-7.5-1202 24-00-T1	1	1.17	8.40	9.57	9.06	93.9	
P5095-01	MH-764	2	1.15	8.81	9.96	8.93	88.3	
P5095-02	MH-764-EPH	3	1.16	8.44	9.6	8.56	87.7	
P5095-03	MH-764-VOC	4	1.14	8.74	9.88	8.89	88.7	
P5096-01	MH-B	5	1.12	8.78	9.9	8.67	86.0	
P5096-02	MH-B-EPH	6	1.15	8.83	9.98	8.94	88.2	
P5096-03	MH-B-VOC	7	1.12	8.85	9.97	9.5	94.7	
P5096-05	MH-A	8	1.14	8.82	9.96	8.9	88.0	
P5096-06	MH-A-EPH	9	1.19	8.52	9.71	8.42	84.9	
P5096-07	MH-A-VOC	10	1.16	8.83	9.99	8.88	87.4	
P5098-01	TR-04-12042024	33	1.16	8.46	9.62	8.87	91.1	
P5098-02	TR-04-12042024	34	1.15	8.53	9.68	8.9	90.9	
P5099-01	BC274770-1-1	11	1.00	1.00	2.00	2.00	100.0	oilc
P5099-02	BC274770-1-2	12	1.00	1.00	2.00	2.00	100.0	oilc
P5100-01	324	13	1.15	8.52	9.67	9.02	92.4	
P5100-02	324-E2	14	1.12	8.68	9.8	9.32	94.5	
P5100-04	3167	15	1.00	1.00	2.00	2.00	100.0	oil sample
P5101-01	BC151973-1-1	16	1.00	1.00	2.00	2.00	100.0	oilc
P5101-02	BC151973-1-2	17	1.00	1.00	2.00	2.00	100.0	oilc
P5101-03	Y2404-0012-1-1	18	1.00	1.00	2.00	2.00	100.0	oilc
P5101-04	Y2404-0012-1-2	19	1.00	1.00	2.00	2.00	100.0	oilc
P5101-05	Y2404-0012-2-1	20	1.00	1.00	2.00	2.00	100.0	oilc
P5101-06	Y2404-0012-2-2	21	1.00	1.00	2.00	2.00	100.0	oilc
P5101-07	HZH012P-1-1	22	1.00	1.00	2.00	2.00	100.0	oilc
P5101-08	HZH012P-1-2	23	1.00	1.00	2.00	2.00	100.0	oilc
P5101-09	Y2404-0034-1-1	24	1.00	1.00	2.00	2.00	100.0	oilc
P5101-10	Y2404-0034-1-2	25	1.00	1.00	2.00	2.00	100.0	oilc
P5102-01	HIH3701-1-1	26	1.00	1.00	2.00	2.00	100.0	oilc

PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/5/2024

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:30  
In Date: 12/04/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:14  
Out Date: 12/05/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB133737

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P5102-02	HIH3701-1-2	27	1.00	1.00	2.00	2.00	100.0	oilc
P5102-03	BC247637-1-1	28	1.00	1.00	2.00	2.00	100.0	oilc
P5102-04	BC247637-1-2	29	1.00	1.00	2.00	2.00	100.0	oilc
P5102-05	BC247637-2-1	30	1.00	1.00	2.00	2.00	100.0	oilc
P5102-06	BC247637-2-2	31	1.00	1.00	2.00	2.00	100.0	oilc
P5103-02	423	32	1.00	1.00	2.00	2.00	100.0	oil sample
P5105-01	CTWK-COMP-1	35	1.14	8.80	9.94	8.55	84.2	
P5105-02	CTWK-COMP-1-E2	36	1.19	8.50	9.69	8.03	80.5	
P5108-01	ASPHALT-COMP	37	1.00	1.00	2.00	2.00	100.0	asphalt sample
P5108-02	ASPHALT-COMP-E2	38	1.00	1.00	2.00	2.00	100.0	asphalt sample
P5110-01	ELIZ-COMP-1	39	1.16	8.44	9.6	8.96	92.4	
P5110-02	ELIZ-COMP-2	40	1.00	1.00	2.00	2.00	100.0	debris

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

133731

WorkList Name : %1-120424

WorkList ID : 185937

Department : Wet-Chemistry

Date : 12-04-2024 08:04:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5076-01	TAPIAL2-SB02I-7.5-120224-00-	Solid	Percent Solids	Cool 4 deg C	WEST04	L61	12/02/2024	Chemtech -SO
P5095-01	MH-764	Solid	Percent Solids	Cool 4 deg C	PSEG03	L11	12/04/2024	Chemtech -SO
P5095-02	MH-764-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L11	12/04/2024	Chemtech -SO
P5095-03	MH-764-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L11	12/04/2024	Chemtech -SO
P5096-01	MH-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5096-02	MH-B-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5096-03	MH-B-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5096-05	MH-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5096-06	MH-A-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5096-07	MH-A-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5098-01	TR-04-12042024	Solid	Percent Solids	Cool 4 deg C	PSEG05	L51	12/04/2024	Chemtech -SO
P5098-02	TR-04-12042024	Solid	Percent Solids	Cool 4 deg C	PSEG05	L51	12/04/2024	Chemtech -SO
P5099-01	BC274770-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5099-02	BC274770-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5100-01	324	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5100-02	324-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5100-04	3167	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5101-01	BC151973-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-02	BC151973-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-03	Y2404-0012-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-04	Y2404-0012-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO

Date/Time 12/04/24 16:25

Raw Sample Received by: JO GPC

Raw Sample Relinquished by: R3 CEA - (ah)



## WORKLIST(Hardcopy Internal Chain)

133737

WorkList Name : %1-120424

WorkList ID : 185937

Department : Wet-Chemistry

Date : 12-04-2024 08:04:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5101-05	Y2404-0012-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-06	Y2404-0012-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-07	HZH012P-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-08	HZH012P-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-09	Y2404-0034-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5101-10	Y2404-0034-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5102-01	HIH3701-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5102-02	HIH3701-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5102-03	BC247637-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5102-04	BC247637-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5102-05	BC247637-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5102-06	BC247637-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5103-02	423	Solid	Percent Solids	Cool 4 deg C	PSEG03	L61	12/04/2024	Chemtech -SO
P5105-01	CTWK-COMP-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	12/04/2024	Chemtech -SO
P5105-02	CTWK-COMP-1-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	12/04/2024	Chemtech -SO
P5108-01	ASPHALT-COMP	Solid	Percent Solids	Cool 4 deg C	PSEG03	L21	12/04/2024	Chemtech -SO
P5108-02	ASPHALT-COMP-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	12/04/2024	Chemtech -SO
P5110-01	ELIZ-COMP-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	12/04/2024	Chemtech -SO
P5110-02	ELIZ-COMP-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	12/04/2024	Chemtech -SO

Date/Time

12/04/24 16:25

Raw Sample Received by:

JH WOC

Raw Sample Relinquished by:

RS Ext. Lab

Date/Time

12/04/24

Raw Sample Received by:

RCEA - Lab

Raw Sample Relinquished by:

JH WOC





# SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

P5076



Weston COC ID
Weston_20241202_1547

### Chain of Custody Record/Lab Work Request

<b>Client:</b>	Weston Solutions, Inc.		
<b>Project Manager:</b>	David Sembrot		
<b>Street Address:</b>	1400 Weston Way	<b>City:</b>	West Chester
<b>Phone:</b>	610-314-5456	<b>ST, ZIP:</b>	PA, 19038
<b>e-mail:</b>	<a href="mailto:david.sembrot@westonsolutions.com">david.sembrot@westonsolutions.com</a>		
<b>Sampled By:</b>	Cheyenne Harrington		

<b>Project Name:</b>	Fort Meade RI	<b>Project POC:</b>	Nathan Fretz
<b>PO Number</b>	0111169	<b>Phone:</b>	484-524-5665
<b>W.O. #:</b>		<b>POC e-mail:</b>	<a href="mailto:nathan.fretz@westonsolutions.com">nathan.fretz@westonsolutions.com</a>
<b>Lab:</b>	CHEMTECH	<b>Lab POC:</b>	Jordan Hedvat
<b>TAT (days):</b>	21	<b>Lab Phone:</b>	908-728-3144
<b>Lab Address:</b>	284 Sheffield Street Mountainside, NJ 07092		

Matrix Codes
SB- Soil
SE - Sediment
SO - Solid
SL - Sludge
GW - Groundwater
W - Water
O - Oil
A - Air
DS - Drum Solids
DL - Drum Liquids
L - EP/TCLP Leachate
WI - Wipe
X - Other
F - Fish

Lab Use Only		
Temperature of cooler when received (°C)		
COC Tape was present and unbroken on outer package?	Y	N
Samples received in good condition?	Y	N
Labels indicate properly preserved?	Y	N
Received within holding times?	Y	N
Discrepancies between sample labels and COC record?	Y	N

<b>Analyses Requested:</b>	pH by EPA 9045D	TAL Metals by EPA 6020B/7471B	TOC by 9060A															
	<b>Container Type:</b>	Glass	Glass	Glass														
	<b>Container Size:</b>	8 oz	8 oz	8 oz														
	<b>Preservative:</b>	Ice to 0-6	Ice to 0-6	Ice to 0-6														

#	Sample ID	G/C	Matrix	# Cont	MS/MSD	Date Collected	Time Collected												Special Instructions/Comments
1	TAPIAL2-SB02I-7.5-120224-00-T1	g	SB	2	no	12/2/2024	15:00	X	X	X									
2																			
3																			
4																			
5																			
6																			
7																			
8																			
9																			
10																			
11																			
12																			

<b>Shipping Airbill Number(s):</b> 7704 1914 8491						<b>Cooler Number:</b> 1 of 1			
	<b>Relinquished By</b>	<b>Date</b>	<b>Time</b>	<b>Received By</b>	<b>Date</b>	<b>Time</b>	<b>Additional Comments</b>		
1.)	<i>[Signature]</i>	12/2/24	1800	<i>[Signature]</i>	12-5-24	10:20	QSM 6.0 Compliant		
2.)							Deliverable Requirements: DoD Level IV report, EnviroData EDD, and ERIS-compatible EDD		
3.)									

*A. Gunkel 3.4.24*

**Laboratory Certification**

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488