

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5076

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133750

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	12/05/2024
Sample ID: CCV1 pH	pH	2.02	2.00	101	90-110	12/05/2024
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	12/05/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5076

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133779

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	934	1000	93	90-110	11/12/2024
Sample ID:	CCV1						
TOC		mg/L	937	1000	94	90-110	12/06/2024
Sample ID:	CCV2						
TOC		mg/L	999	1000	100	90-110	12/06/2024