

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51879MA No.: _____ SDG No.: MBHLL1Instrument ID: P5Start Date : 12/27/2024Analytical Method: ICP-AESRun Batch : LB134114

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999842	0.100060	-0.000266	WLR Blank	Inverse Conc
Antimony	0.995864	0.118106	0.000254	WLR Blank	Inverse Conc
Arsenic	0.999800	156.923530	0.027719	WLR Blank	Inverse Conc
Barium	0.999930	3.573707	0.091981	WLR Blank	Inverse Conc
Beryllium	0.999966	10.047430	0.001742	WLR Blank	Inverse Conc
Boron	0.999556	0.945534	0.007337	WLR Blank	Inverse Conc
Cadmium	0.999923	5.725509	0.000361	WLR Blank	Inverse Conc
Calcium	0.999979	0.073276	0.003639	WLR Blank	Inverse Conc
Chromium	0.999973	0.193839	0.000203	WLR Blank	Inverse Conc
Cobalt	0.999920	1.017961	-0.000304	WLR Blank	Inverse Conc
Copper	0.999964	0.197422	0.002559	WLR Blank	Inverse Conc
Iron	0.999957	0.021475	-0.000278	WLR Blank	Inverse Conc
Lead	0.999756	0.360619	0.000383	WLR Blank	Inverse Conc
Lithium	0.999899	1.101915	-0.003424	WLR Blank	Inverse Conc
Magnesium	0.999979	0.010188	-0.001401	WLR Blank	Inverse Conc
Manganese	0.999960	0.476790	0.000353	WLR Blank	Inverse Conc
Molybdenum	0.996721	0.589536	0.000217	WLR Blank	Inverse Conc
Nickel	0.999885	0.953359	-0.000284	WLR Blank	Inverse Conc
Phosphorus	0.999370	0.104279	-0.002554	WLR Blank	Inverse Conc
Potassium	0.999691	0.026202	0.078852	WLR Blank	Inverse Conc
Selenium	0.997621	0.074303	-0.000085	WLR Blank	Inverse Conc
Silicon	0.994603	1.429473	0.014317	WLR Blank	Inverse Conc
Silver	0.999784	0.260195	-0.000392	WLR Blank	Inverse Conc
Sodium	0.998153	0.376774	-0.031962	WLR Blank	Inverse Conc
Strontium	0.999965	7.164940	-0.002159	WLR Blank	Inverse Conc
Sulfur	0.999045	0.038810	0.000345	WLR Blank	Inverse Conc
Thallium	0.999974	0.127996	-0.000223	WLR Blank	Inverse Conc
Tin	0.999736	0.147576	0.000613	WLR Blank	Inverse Conc
Vanadium	0.999953	0.102531	0.000187	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHLL1
Instrument ID: P5 Start Date : 12/27/2024
Analytical Method: ICP-AES Run Batch : LB134114

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.997774	1.739977	0.004113	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51879MA No.: _____ SDG No.: MBHLL1Instrument ID: P5Start Date : 01/03/2025Analytical Method: ICP-AESRun Batch : LB134161

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999769	0.103654	-0.001819	WLR Blank	Inverse Conc
Antimony	0.998614	0.169726	-0.000032	WLR Blank	Inverse Conc
Arsenic	0.999767	131.619340	0.189867	WLR Blank	Inverse Conc
Barium	0.999935	3.393608	0.167881	WLR Blank	Inverse Conc
Beryllium	0.999935	13.524085	0.004962	WLR Blank	Inverse Conc
Boron	0.999460	1.127550	0.008656	WLR Blank	Inverse Conc
Cadmium	0.999929	4.450103	0.000501	WLR Blank	Inverse Conc
Calcium	0.999994	0.054802	0.011547	WLR Blank	Inverse Conc
Chromium	0.999995	0.351633	-0.000110	WLR Blank	Inverse Conc
Cobalt	0.999965	1.167159	-0.000821	WLR Blank	Inverse Conc
Copper	0.999962	0.151250	0.003132	WLR Blank	Inverse Conc
Iron	0.999987	0.024081	-0.000208	WLR Blank	Inverse Conc
Lead	0.999791	0.286633	-0.000676	WLR Blank	Inverse Conc
Lithium	0.999899	1.422611	-0.001239	WLR Blank	Inverse Conc
Magnesium	0.999988	0.011958	-0.000058	WLR Blank	Inverse Conc
Manganese	0.999996	0.449550	0.000088	WLR Blank	Inverse Conc
Molybdenum	0.998629	0.626497	0.001115	WLR Blank	Inverse Conc
Nickel	0.999911	0.694264	-0.001274	WLR Blank	Inverse Conc
Phosphorus	0.999611	0.093497	-0.004519	WLR Blank	Inverse Conc
Potassium	0.999631	0.032523	0.066515	WLR Blank	Inverse Conc
Selenium	0.999349	0.070441	-0.000127	WLR Blank	Inverse Conc
Silicon	0.994217	2.283946	0.022974	WLR Blank	Inverse Conc
Silver	0.999853	0.333831	0.000171	WLR Blank	Inverse Conc
Sodium	0.998008	0.682872	0.102999	WLR Blank	Inverse Conc
Strontium	0.999900	9.493908	-0.001853	WLR Blank	Inverse Conc
Sulfur	0.999184	0.055803	0.001153	WLR Blank	Inverse Conc
Thallium	0.999976	0.095659	-0.000448	WLR Blank	Inverse Conc
Tin	0.999751	0.141695	0.002549	WLR Blank	Inverse Conc
Vanadium	0.999892	0.108400	-0.000235	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHLL1
Instrument ID: P5 Start Date : 01/03/2025
Analytical Method: ICP-AES Run Batch : LB134161

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.999297	2.741170	0.002813	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHLL1
Instrument ID: P5 Start Date : 01/03/2025
Analytical Method: ICP-AES Run Batch : LB134161

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------