

Surrogate Summary

SDG No.: **P5088**

Client: **Weston Solutions, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5088-01	TB-241203-01	1,2-Dichloroethane-d4	50	53.2	106	70 (74)	130 (125)
		Dibromofluoromethane	50	51.3	103	70 (75)	130 (124)
		Toluene-d8	50	48.2	96	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.7	87	70 (77)	130 (121)
P5088-02	UNIT-51-FS-SW001-01	1,2-Dichloroethane-d4	50	52.5	105	70 (74)	130 (125)
		Dibromofluoromethane	50	50.8	102	70 (75)	130 (124)
		Toluene-d8	50	47.5	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.0	86	70 (77)	130 (121)
P5088-03	UNIT-51-RS-SW001-01	1,2-Dichloroethane-d4	50	52.5	105	70 (74)	130 (125)
		Dibromofluoromethane	50	49.7	99	70 (75)	130 (124)
		Toluene-d8	50	47.1	94	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.1	86	70 (77)	130 (121)
P5088-04MS	UNIT-51-RS-SW001-01MS	1,2-Dichloroethane-d4	50	51.0	102	70 (74)	130 (125)
		Dibromofluoromethane	50	54.2	108	70 (75)	130 (124)
		Toluene-d8	50	52.6	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.0	106	70 (77)	130 (121)
P5088-05MSD	UNIT-51-RS-SW001-01MSD	1,2-Dichloroethane-d4	50	49.4	99	70 (74)	130 (125)
		Dibromofluoromethane	50	51.1	102	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.2	102	70 (77)	130 (121)
P5088-06	UNIT-51-RS-SW001-02	1,2-Dichloroethane-d4	50	52.5	105	70 (74)	130 (125)
		Dibromofluoromethane	50	50.1	100	70 (75)	130 (124)
		Toluene-d8	50	46.4	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	40.9	82	70 (77)	130 (121)
VN1204WBL01	VN1204WBL01	1,2-Dichloroethane-d4	50	51.3	103	70 (74)	130 (125)
		Dibromofluoromethane	50	51.0	102	70 (75)	130 (124)
		Toluene-d8	50	47.3	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.2	86	70 (77)	130 (121)
VN1204WBS01	VN1204WBS01	1,2-Dichloroethane-d4	50	48.3	97	70 (74)	130 (125)
		Dibromofluoromethane	50	50.9	102	70 (75)	130 (124)
		Toluene-d8	50	49.6	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.1	96	70 (77)	130 (121)
VN1205WBL01	VN1205WBL01	1,2-Dichloroethane-d4	50	51.7	103	70 (74)	130 (125)
		Dibromofluoromethane	50	50.5	101	70 (75)	130 (124)
		Toluene-d8	50	48.3	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.0	88	70 (77)	130 (121)
VN1205WBS01	VN1205WBS01	1,2-Dichloroethane-d4	50	50.8	102	70 (74)	130 (125)
		Dibromofluoromethane	50	53.1	106	70 (75)	130 (124)
		Toluene-d8	50	50.8	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.2	100	70 (77)	130 (121)

