

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5088

Client: Weston Solutions, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec		RPD	Limits			
		Result			Qual	RPD	Qual	Low	High	RPD	
Lab Sample ID :	P5088-04MS	Client Sample ID :	UNIT-51-RS-SW001-01MS					Datafile :	VN085102.D		
Vinyl chloride	50	0	42.5	ug/L	85				70 (69)	130 (125)	
1,1-Dichloroethene	50	0	46.8	ug/L	94				70 (53)	130 (162)	
Methylene Chloride	50	0	46.6	ug/L	93				70 (79)	130 (115)	
trans-1,2-Dichloroethene	50	0	45.2	ug/L	90				70 (76)	130 (118)	
1,1-Dichloroethane	50	0	49.2	ug/L	98				70 (78)	130 (122)	
Carbon Tetrachloride	50	0	52.2	ug/L	104				70 (66)	130 (133)	
cis-1,2-Dichloroethene	50	0	50.0	ug/L	100				70 (82)	130 (124)	
1,1,1-Trichloroethane	50	0	50.1	ug/L	100				70 (83)	130 (117)	
Trichloroethene	50	0.76	50.5	ug/L	99				70 (28)	130 (175)	
1,1,2-Trichloroethane	50	0	56.5	ug/L	113				70 (27)	130 (175)	
Tetrachloroethene	50	0	47.5	ug/L	95				70 (48)	130 (153)	
1,2,3-Trichloropropane	50	0	52.8	ug/L	106				70 (68)	130 (128)	