

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5088

Client: Weston Solutions, Inc.

Analytical Method: SW8260-Low

Datafile : VN085109.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1205WBS01	Vinyl chloride	20	18.9	ug/L	95			70 (65)	130 (117)	
	1,1-Dichloroethene	20	20.7	ug/L	104			70 (74)	130 (110)	
	Methylene Chloride	20	20.1	ug/L	101			70 (72)	130 (114)	
	trans-1,2-Dichloroethene	20	20.0	ug/L	100			70 (75)	130 (108)	
	1,1-Dichloroethane	20	20.5	ug/L	103			70 (78)	130 (112)	
	Carbon Tetrachloride	20	22.1	ug/L	111			70 (77)	130 (113)	
	cis-1,2-Dichloroethene	20	20.8	ug/L	104			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	21.6	ug/L	108			70 (80)	130 (108)	
	Trichloroethene	20	21.8	ug/L	109			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	23.8	ug/L	119			70 (83)	130 (112)	
	Tetrachloroethene	20	24.2	ug/L	121			70 (67)	130 (123)	
	1,2,3-Trichloropropane	20	23.7	ug/L	119			70 (75)	130 (112)	