

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ROYF02
 Lab Code: CHEM Case No.: P5088 SAS No.: P5088 SDG NO.: P5088
 Lab File ID: VN085106.D Date Analyzed: 12/05/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	121642	13.788				
UPPER LIMIT	243284	14.288				
LOWER LIMIT	60821	13.288				
EPA SAMPLE NO.						
TB-241203-01	89088	13.79				
VN1205WBL01	93689	13.79				
VN1205WBS01	104397	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.