

## Initial and Continuing Calibration Verification

**Client:** R3M Engineering, Inc.

**SDG No.:** P5093

**Project:** MC054.36 23-8-1 LM - Landing Lane New Brunswick

**RunNo.:** LB133751

| Analyte                       | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|-------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>TOC | mg/L  | 10.2   | 10         | 102        | 90-110                 | 09/25/2024    |
| Sample ID: <b>CCV1</b><br>TOC | mg/L  | 10.7   | 10         | 107        | 90-110                 | 12/05/2024    |
| Sample ID: <b>CCV2</b><br>TOC | mg/L  | 10.9   | 10         | 109        | 90-110                 | 12/05/2024    |
| Sample ID: <b>CCV3</b><br>TOC | mg/L  | 10.6   | 10         | 106        | 90-110                 | 12/05/2024    |

## Initial and Continuing Calibration Verification

**Client:** R3M Engineering, Inc.

**SDG No.:** P5093

**Project:** MC054.36 23-8-1 LM - Landing Lane New Brunswick

**RunNo.:** LB133847

| Analyte    |      | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------|------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: | ICV1 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.1    | 0.099      | 101        | 90-110                 | 12/09/2024    |
| Sample ID: | CCV1 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.26   | 0.25       | 104        | 90-110                 | 12/09/2024    |
| Sample ID: | CCV2 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.25   | 0.25       | 100        | 90-110                 | 12/09/2024    |
| Sample ID: | CCV3 |       |        |            |            |                        |               |
| Cyanide    |      | mg/L  | 0.27   | 0.25       | 108        | 90-110                 | 12/09/2024    |