

Initial and Continuing Calibration Blank Summary

Client: R3M Engineering, Inc.

SDG No.: P5093

Project: MC054.36 23-8-1 LM - Landing Lane New Brunswick

RunNo.: LB133751

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	< 0.5000	0.5000	U	0.19	1	09/25/2024
Sample ID: CCB1 TOC	mg/L	0.43	0.5000	J	0.19	1	12/05/2024
Sample ID: CCB2 TOC	mg/L	0.39	0.5000	J	0.19	1	12/05/2024
Sample ID: CCB3 TOC	mg/L	0.23	0.5000	J	0.19	1	12/05/2024

Initial and Continuing Calibration Blank Summary

Client: R3M Engineering, Inc.

SDG No.: P5093

Project: MC054.36 23-8-1 LM - Landing Lane New Brunswick

RunNo.: LB133847

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024