

CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

Continuing Calib Date: 12/06/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 09:50 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.93	5.92	5.82	6.02	-0.01
Aroclor-1242-2 (2)	5.95	5.95	5.85	6.05	0.00
Aroclor-1242-3 (3)	6.01	6.01	5.91	6.11	0.00
Aroclor-1242-4 (4)	6.11	6.11	6.01	6.21	0.00
Aroclor-1242-5 (5)	6.85	6.84	6.74	6.94	0.00
Tetrachloro-m-xylene	4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl	10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

Continuing Calib Date: 12/06/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 09:50 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1242-2 (2)	5.18	5.18	5.08	5.28	0.00
Aroclor-1242-3 (3)	5.36	5.36	5.26	5.46	0.00
Aroclor-1242-4 (4)	5.44	5.44	5.34	5.54	0.00
Aroclor-1242-5 (5)	5.97	5.97	5.87	6.07	0.00
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.20	9.21	9.11	9.31	0.01