

### CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

Continuing Calib Date: 12/06/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 10:06 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.93       | 5.92      | 5.82      | 6.02  | -0.01      |
| Aroclor-1248-2 (2)   | 6.20       | 6.19      | 6.09      | 6.29  | -0.01      |
| Aroclor-1248-3 (3)   | 6.40       | 6.40      | 6.30      | 6.50  | 0.00       |
| Aroclor-1248-4 (4)   | 6.80       | 6.80      | 6.70      | 6.90  | 0.00       |
| Aroclor-1248-5 (5)   | 6.84       | 6.84      | 6.74      | 6.94  | 0.00       |
| Tetrachloro-m-xylene | 4.76       | 4.76      | 4.66      | 4.86  | 0.00       |
| Decachlorobiphenyl   | 10.68      | 10.67     | 10.57     | 10.77 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

Continuing Calib Date: 12/06/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 10:06 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1248-2 (2)   | 5.39       | 5.40      | 5.30      | 5.50 | 0.01       |
| Aroclor-1248-3 (3)   | 5.44       | 5.44      | 5.34      | 5.54 | 0.00       |
| Aroclor-1248-4 (4)   | 5.61       | 5.62      | 5.52      | 5.72 | 0.01       |
| Aroclor-1248-5 (5)   | 6.01       | 6.01      | 5.91      | 6.11 | 0.00       |
| Tetrachloro-m-xylene | 4.05       | 4.05      | 3.95      | 4.15 | 0.00       |
| Decachlorobiphenyl   | 9.20       | 9.21      | 9.11      | 9.31 | 0.01       |