

CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

Continuing Calib Date: 12/06/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 10:23 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.78	6.78	6.68	6.88	0.00
Aroclor-1254-2 (2)	7.00	7.00	6.90	7.10	0.00
Aroclor-1254-3 (3)	7.37	7.36	7.26	7.46	0.00
Aroclor-1254-4 (4)	7.65	7.64	7.54	7.74	-0.01
Aroclor-1254-5 (5)	8.07	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl	10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: RTHR01

Lab Code: CHEM Case No.: P5093 SAS No.: P5093 SDG NO.: P5093

Continuing Calib Date: 12/06/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 10:23 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.97	5.97	5.87	6.07	0.00
Aroclor-1254-2 (2)	6.12	6.12	6.02	6.22	0.01
Aroclor-1254-3 (3)	6.52	6.53	6.43	6.63	0.01
Aroclor-1254-4 (4)	6.75	6.76	6.66	6.86	0.01
Aroclor-1254-5 (5)	7.17	7.18	7.08	7.28	0.01
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.20	9.21	9.11	9.31	0.01